

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB042822\
 Data File : FB027101.D
 Signal(s) : FID2B.CH
 Acq On : 28 Apr 2022 13:12
 Operator : AJ/MA
 Sample : BSF0428S2
 Misc : 5.03G/5.00 ML DI WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF0428S2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/13/2022
 Supervised By :mohammad ahmed 05/13/2022

Integration File: Calibration.e
 Quant Time: Apr 29 02:23:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040922.M
 Quant Title :
 QLast Update : Sat Apr 09 04:06:25 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.790	226950	20.074 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.707	436494	23.189 ng/ml
2) t 2,2,4-Trimethylpentane	7.420	501104	25.048 ng/ml
3) t n-Heptane	7.750	151347	7.737 ng/ml
4) t Benzene	7.884	200229	9.073 ng/ml
6) t Toluene	10.617	579197	28.089 ng/ml
7) t Ethylbenzene	13.054	186204	9.745 ng/ml
8) t m-Xylene	13.186	353956	18.738 ng/ml
9) t O-Xylene	13.913	340188	18.781 ng/ml
10) t 1,2,4-Trimethylbenzene	16.186	469190	25.456 ng/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB042822\
Data File : FB027101.D
Signal(s) : FID2B.CH
Acq On : 28 Apr 2022 13:12
Operator : AJ/MA
Sample : BSF0428S2
Misc : 5.03G/5.00 ML DI WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF0428S2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/13/2022
Supervised By :mohammad ahmed 05/13/2022

Integration File: Calibration.e
Quant Time: Apr 29 02:23:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB040922.M
Quant Title :
QLast Update : Sat Apr 09 04:06:25 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

