

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB043019\
 Data File : FB020869.D
 Signal(s) : FID2B.CH
 Acq On : 30 Apr 2019 9:55
 Operator : MJ/SJ
 Sample : BSF0430S1
 Misc : 5.02g/5.00ml DI WATER
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: May 01 03:32:35 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB042919.M
 Quant Title :
 QLast Update : Mon Apr 29 15:53:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.994	343969	19.174 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.857	1366852	30.301 ng/ml
2) t 2,2,4-Trimethylpentane	7.609	1287919	30.897 ng/ml
3) t n-Heptane	7.950	405029	10.231 ng/ml
4) t Benzene	8.086	421443	10.074 ng/ml
6) t Toluene	10.838	839143	30.292 ng/ml
7) t Ethylbenzene	13.289	272110	9.702 ng/ml
8) t m-Xylene	13.422	580948	19.547 ng/ml
9) t O-Xylene	14.153	604843	19.513 ng/ml
10) t 1,2,4-Trimethylbenzene	16.436	612046	19.735 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB043019\
Data File : FB020869.D
Signal(s) : FID2B.CH
Acq On : 30 Apr 2019 9:55
Operator : MJ/SJ
Sample : BSF0430S1
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: May 01 03:32:35 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB042919.M
Quant Title :
QLast Update : Mon Apr 29 15:53:58 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.857 min
Delta R.T.: 0.000 min
Response: 1366852
Conc: 30.30 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.609 min
Delta R.T.: 0.000 min
Response: 1287919
Conc: 30.90 ng/ml




#3 n-Heptane

R.T.: 7.950 min
Delta R.T.: 0.001 min
Response: 405029
Conc: 10.23 ng/ml




#4 Benzene

R.T.: 8.086 min
Delta R.T.: 0.002 min
Response: 421443
Conc: 10.07 ng/ml




#5 AAA-TFT

R.T.: 8.994 min
Delta R.T.: 0.002 min
Response: 343969
Conc: 19.17 ng/ml




#6 Toluene

R.T.: 10.838 min
Delta R.T.: 0.002 min
Response: 839143
Conc: 30.29 ng/ml




#7 Ethylbenzene

R.T.: 13.289 min
Delta R.T.: 0.003 min
Response: 272110
Conc: 9.70 ng/ml




#8 m-Xylene

R.T.: 13.422 min
Delta R.T.: 0.003 min
Response: 580948
Conc: 19.55 ng/ml



#9 O-Xylene

R.T.: 14.153 min
Delta R.T.: 0.003 min
Response: 604843
Conc: 19.51 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.436 min
Delta R.T.: 0.003 min
Response: 612046
Conc: 19.73 ng/ml