

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB050219\
 Data File : FB020920.D
 Signal(s) : FID2B.CH
 Acq On : 2 May 2019 8:43
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: May 03 01:39:09 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB042919.M
 Quant Title :
 QLast Update : Mon Apr 29 15:53:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.991	345621	19.266 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.855	1172726	25.998 ng/ml
2) t 2,2,4-Trimethylpentane	7.608	1088018	26.102 ng/ml
3) t n-Heptane	7.947	344453	8.701 ng/ml
4) t Benzene	8.083	376027	8.988 ng/ml
6) t Toluene	10.836	764564	27.600 ng/ml
7) t Ethylbenzene	13.286	270578	9.648 ng/ml
8) t m-Xylene	13.419	585060	19.685 ng/ml
9) t O-Xylene	14.150	601050	19.391 ng/ml
10) t 1,2,4-Trimethylbenzene	16.434	616360	19.874 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB050219\
Data File : FB020920.D
Signal(s) : FID2B.CH
Acq On : 2 May 2019 8:43
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: May 03 01:39:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB042919.M
Quant Title :
QLast Update : Mon Apr 29 15:53:58 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.855 min
Delta R.T.: -0.002 min
Response: 1172726
Conc: 26.00 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.608 min
Delta R.T.: -0.001 min
Response: 1088018
Conc: 26.10 ng/ml




#3 n-Heptane

R.T.: 7.947 min
Delta R.T.: -0.001 min
Response: 344453
Conc: 8.70 ng/ml




#4 Benzene


R.T.: 8.083 min
Delta R.T.: 0.000 min
Response: 376027
Conc: 8.99 ng/ml




#5 AAA-TFT
 R.T.: 8.991 min
 Delta R.T.: 0.000 min
 Response: 345621
 Conc: 19.27 ng/ml




#6 Toluene
 R.T.: 10.836 min
 Delta R.T.: 0.000 min
 Response: 764564
 Conc: 27.60 ng/ml



#7 Ethylbenzene
 R.T.: 13.286 min
 Delta R.T.: 0.000 min
 Response: 270578
 Conc: 9.65 ng/ml




#8 m-Xylene
 R.T.: 13.419 min
 Delta R.T.: 0.000 min
 Response: 585060
 Conc: 19.69 ng/ml



#9 O-Xylene

R.T.: 14.150 min
Delta R.T.: 0.000 min
Response: 601050
Conc: 19.39 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.434 min
Delta R.T.: 0.000 min
Response: 616360
Conc: 19.87 ng/ml