

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB050219\
 Data File : FB020941.D
 Signal(s) : FID2B.CH
 Acq On : 2 May 2019 22:51
 Operator : MJ/SJ
 Sample : K2672-18
 Misc : 5.03g/5.00ml DI WATER
 ALS Vial : 22 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: May 03 01:42:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB042919.M
 Quant Title :
 QLast Update : Mon Apr 29 15:53:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.993	303868	16.939 ng/ml
Target Compounds			
7) t Ethylbenzene	13.290	5917	0.211 ng/ml
8) t m-Xylene	13.416	19469	0.655 ng/ml
9) t O-Xylene	14.155	10364	0.334 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB050219\
Data File : FB020941.D
Signal(s) : FID2B.CH
Acq On : 2 May 2019 22:51
Operator : MJ/SJ
Sample : K2672-18
Misc : 5.03g/5.00ml DI WATER
ALS Vial : 22 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: May 03 01:42:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB042919.M
Quant Title :
QLast Update : Mon Apr 29 15:53:58 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#5 AAA-TFT

R.T.: 8.993 min
Delta R.T.: 0.001 min
Response: 303868
Conc: 16.94 ng/ml




#7 Ethylbenzene

R.T.: 13.290 min
Delta R.T.: 0.004 min
Response: 5917
Conc: 0.21 ng/ml



#8 m-Xylene

R.T.: 13.416 min
Delta R.T.: -0.003 min
Response: 19469
Conc: 0.66 ng/ml



#9 O-Xylene

R.T.: 14.155 min
Delta R.T.: 0.006 min
Response: 10364
Conc: 0.33 ng/ml