

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB050521\
 Data File : FB025699.D
 Signal(s) : FID2B.CH
 Acq On : 5 May 2021 14:36
 Operator : DD/AJ
 Sample : 100 GRO STD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: May 06 14:57:46 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB050521.M
 Quant Title :
 QLast Update : Thu May 06 14:51:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.189	2153929	102.091 ng/ml
Target Compounds			
2) t 2,2,4-Trimethylpentane	7.825	3202270	161.405 ng/ml
3) t n-Heptane	8.148	1676101	59.370 ng/ml
4) t Benzene	8.298	2027764	54.296 ng/ml
6) t Toluene	11.026	5855904	156.867 ng/ml
7) t Ethylbenzene	13.449	1886301	53.182 ng/ml
8) t m-Xylene	13.580	3767621	105.840 ng/ml
9) t O-Xylene	14.306	3607925	111.051 ng/ml
10) t 1,2,4-Trimethylbenzene	16.565	2934062	106.793 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB050521\
Data File : FB025699.D
Signal(s) : FID2B.CH
Acq On : 5 May 2021 14:36
Operator : DD/AJ
Sample : 100 GRO STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: May 06 14:57:46 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB050521.M
Quant Title :
QLast Update : Thu May 06 14:51:17 2021
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#2 2,2,4-Trimethylpentane

R.T.: 7.825 min
Delta R.T.: -0.058 min
Response: 3202270
Conc: 161.41 ng/ml




#3 n-Heptane

R.T.: 8.148 min
Delta R.T.: -0.054 min
Response: 1676101
Conc: 59.37 ng/ml




#4 Benzene

R.T.: 8.298 min
Delta R.T.: -0.057 min
Response: 2027764
Conc: 54.30 ng/ml




#5 AAA-TFT

R.T.: 9.189 min
Delta R.T.: -0.041 min
Response: 2153929
Conc: 102.09 ng/ml




#6 Toluene

R.T.: 11.026 min
Delta R.T.: -0.026 min
Response: 5855904
Conc: 156.87 ng/ml




#7 Ethylbenzene

R.T.: 13.449 min
Delta R.T.: -0.012 min
Response: 1886301
Conc: 53.18 ng/ml




#8 m-Xylene

R.T.: 13.580 min
Delta R.T.: -0.011 min
Response: 3767621
Conc: 105.84 ng/ml



#9 O-Xylene

R.T.: 14.306 min
Delta R.T.: -0.009 min
Response: 3607925
Conc: 111.05 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.565 min
Delta R.T.: -0.003 min
Response: 2934062
Conc: 106.79 ng/ml