

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB050619\
 Data File : FB020973.D
 Signal(s) : FID2B.CH
 Acq On : 6 May 2019 11:48
 Operator : MJ/SJ
 Sample : K2639-03MS
 Misc : 5.03g/5.00ml DI WATER
 ALS Vial : 6 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: May 07 03:14:19 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB042919.M
 Quant Title :
 QLast Update : Mon Apr 29 15:53:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.994	368830	20.560 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.857	969282	21.488 ng/ml
2) t 2,2,4-Trimethylpentane	7.611	866649	20.791 ng/ml
3) t n-Heptane	7.951	264223	6.674 ng/ml
4) t Benzene	8.086	349464	8.353 ng/ml
6) t Toluene	10.839	671917	24.256 ng/ml
7) t Ethylbenzene	13.289	218111	7.777 ng/ml
8) t m-Xylene	13.422	465100	15.649 ng/ml
9) t O-Xylene	14.153	484003	15.615 ng/ml
10) t 1,2,4-Trimethylbenzene	16.437	469565	15.141 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB050619\
Data File : FB020973.D
Signal(s) : FID2B.CH
Acq On : 6 May 2019 11:48
Operator : MJ/SJ
Sample : K2639-03MS
Misc : 5.03g/5.00ml DI WATER
ALS Vial : 6 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: May 07 03:14:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB042919.M
Quant Title :
QLast Update : Mon Apr 29 15:53:58 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.857 min
Delta R.T.: 0.000 min
Response: 969282
Conc: 21.49 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.611 min
Delta R.T.: 0.002 min
Response: 866649
Conc: 20.79 ng/ml




#3 n-Heptane

R.T.: 7.951 min
Delta R.T.: 0.002 min
Response: 264223
Conc: 6.67 ng/ml




#4 Benzene

R.T.: 8.086 min
Delta R.T.: 0.002 min
Response: 349464
Conc: 8.35 ng/ml




#5 AAA-TFT

R.T.: 8.994 min
Delta R.T.: 0.002 min
Response: 368830
Conc: 20.56 ng/ml




#6 Toluene

R.T.: 10.839 min
Delta R.T.: 0.003 min
Response: 671917
Conc: 24.26 ng/ml




#7 Ethylbenzene

R.T.: 13.289 min
Delta R.T.: 0.003 min
Response: 218111
Conc: 7.78 ng/ml




#8 m-Xylene

R.T.: 13.422 min
Delta R.T.: 0.003 min
Response: 465100
Conc: 15.65 ng/ml



#9 O-Xylene

R.T.: 14.153 min
Delta R.T.: 0.003 min
Response: 484003
Conc: 15.61 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.437 min
Delta R.T.: 0.004 min
Response: 469565
Conc: 15.14 ng/ml