

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB051320\
 Data File : FB024353.D
 Signal(s) : FID2B.CH
 Acq On : 13 May 2020 16:53
 Operator : DD/AJ
 Sample : L2182-02
 Misc : 5.01g/5.00ml DI WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: May 14 04:29:35 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051120.M
 Quant Title :
 QLast Update : Mon May 11 11:14:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.799	432727	19.596 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.691	241612	8.187 ng/ml
2) t 2,2,4-Trimethylpentane	7.420	260766	8.536 ng/ml
3) t n-Heptane	7.757	87417	2.885 ng/ml
4) t Benzene	7.892	103676	2.742 ng/ml
6) t Toluene	10.630	295724	8.161 ng/ml
7) t Ethylbenzene	13.072	90520	2.670 ng/ml
8) t m-Xylene	13.203	183984	5.575 ng/ml
9) t O-Xylene	13.932	189480	5.872 ng/ml
10) t 1,2,4-Trimethylbenzene	16.209	147110	5.455 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB051320\
Data File : FB024353.D
Signal(s) : FID2B.CH
Acq On : 13 May 2020 16:53
Operator : DD/AJ
Sample : L2182-02
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: May 14 04:29:35 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051120.M
Quant Title :
QLast Update : Mon May 11 11:14:42 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.691 min
Delta R.T.: 0.002 min
Response: 241612
Conc: 8.19 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.420 min
Delta R.T.: 0.003 min
Response: 260766
Conc: 8.54 ng/ml




#3 n-Heptane

R.T.: 7.757 min
Delta R.T.: -0.001 min
Response: 87417
Conc: 2.89 ng/ml



#4 Benzene


R.T.: 7.892 min
Delta R.T.: 0.000 min
Response: 103676
Conc: 2.74 ng/ml



#5 AAA-TFT


R.T.: 8.799 min
Delta R.T.: 0.000 min
Response: 432727
Conc: 19.60 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.630 min
Delta R.T.: 0.000 min
Response: 295724
Conc: 8.16 ng/ml




#7 Ethylbenzene

R.T.: 13.072 min
Delta R.T.: 0.000 min
Response: 90520
Conc: 2.67 ng/ml



#8 m-Xylene

R.T.: 13.203 min
Delta R.T.: 0.000 min
Response: 183984
Conc: 5.58 ng/ml



#9 O-Xylene


R.T.: 13.932 min

Delta R.T.: 0.000 min

Response: 189480

Conc: 5.87 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.209 min

Delta R.T.: 0.000 min

Response: 147110

Conc: 5.45 ng/ml