

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB051519\
 Data File : FB021041.D
 Signal(s) : FID2B.CH
 Acq On : 15 May 2019 13:47
 Operator : MJ/SJ
 Sample : K2851-02MS
 Misc : 5.03g/5.00ml DI WATER
 ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: May 16 02:23:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB042919.M
 Quant Title :
 QLast Update : Mon Apr 29 15:53:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.998	304079	16.950 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.860	916362	20.315 ng/ml
2) t 2,2,4-Trimethylpentane	7.613	965257	23.157 ng/ml
3) t n-Heptane	7.954	156319	3.949 ng/ml
4) t Benzene	8.091	331786	7.931 ng/ml
6) t Toluene	10.842	570518	20.595 ng/ml
7) t Ethylbenzene	13.293	156779	5.590 ng/ml
8) t m-Xylene	13.426	343405	11.555 ng/ml
9) t O-Xylene	14.157	396380	12.788 ng/ml
10) t 1,2,4-Trimethylbenzene	16.441	357958	11.542 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB051519\
Data File : FB021041.D
Signal(s) : FID2B.CH
Acq On : 15 May 2019 13:47
Operator : MJ/SJ
Sample : K2851-02MS
Misc : 5.03g/5.00ml DI WATER
ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: May 16 02:23:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB042919.M
Quant Title :
QLast Update : Mon Apr 29 15:53:58 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.860 min
Delta R.T.: 0.003 min
Response: 916362
Conc: 20.31 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.613 min
Delta R.T.: 0.004 min
Response: 965257
Conc: 23.16 ng/ml




#3 n-Heptane

R.T.: 7.954 min
Delta R.T.: 0.005 min
Response: 156319
Conc: 3.95 ng/ml




#4 Benzene

R.T.: 8.091 min
Delta R.T.: 0.006 min
Response: 331786
Conc: 7.93 ng/ml




#5 AAA-TFT

R.T.: 8.998 min
Delta R.T.: 0.006 min
Response: 304079
Conc: 16.95 ng/ml




#6 Toluene

R.T.: 10.842 min
Delta R.T.: 0.007 min
Response: 570518
Conc: 20.60 ng/ml




#7 Ethylbenzene

R.T.: 13.293 min
Delta R.T.: 0.007 min
Response: 156779
Conc: 5.59 ng/ml




#8 m-Xylene

R.T.: 13.426 min
Delta R.T.: 0.007 min
Response: 343405
Conc: 11.55 ng/ml



#9 O-Xylene

R.T.: 14.157 min
Delta R.T.: 0.008 min
Response: 396380
Conc: 12.79 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.441 min
Delta R.T.: 0.008 min
Response: 357958
Conc: 11.54 ng/ml