

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB051820\
 Data File : FB024361.D
 Signal(s) : FID2B.CH
 Acq On : 18 May 2020 9:44
 Operator : DD/AJ
 Sample : L2501-02
 Misc : 5.02g/5.00ml DI WATER
 ALS Vial : 4 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: May 18 14:41:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051120.M
 Quant Title :
 QLast Update : Mon May 11 11:14:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.804	631974	28.620 ng/ml
Target Compounds			
2) t 2,2,4-Trimethylpentane	7.383	720	0.024 ng/ml
3) t n-Heptane	7.761	46031	1.519 ng/ml
6) t Toluene	10.630	100751	2.780 ng/ml
7) t Ethylbenzene	13.035	3608143	106.428 ng/ml
8) t m-Xylene	13.185	749466	22.711 ng/ml
9) t O-Xylene	13.928	2709748	83.971 ng/ml
10) t 1,2,4-Trimethylbenzene	16.208	1973527	73.176 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB051820\
Data File : FB024361.D
Signal(s) : FID2B.CH
Acq On : 18 May 2020 9:44
Operator : DD/AJ
Sample : L2501-02
Misc : 5.02g/5.00ml DI WATER
ALS Vial : 4 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: May 18 14:41:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051120.M
Quant Title :
QLast Update : Mon May 11 11:14:42 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#2 2,2,4-Trimethylpentane

R.T.: 7.383 min
Delta R.T.: -0.034 min
Response: 720
Conc: 0.02 ng/ml




#3 n-Heptane

R.T.: 7.761 min
Delta R.T.: 0.003 min
Response: 46031
Conc: 1.52 ng/ml




#5 AAA-TFT

R.T.: 8.804 min
Delta R.T.: 0.004 min
Response: 631974
Conc: 28.62 ng/ml




#6 Toluene

R.T.: 10.630 min
Delta R.T.: 0.000 min
Response: 100751
Conc: 2.78 ng/ml




#7 Ethylbenzene

R.T.: 13.035 min
Delta R.T.: -0.036 min
Response: 3608143
Conc: 106.43 ng/ml




#8 m-Xylene

R.T.: 13.185 min
Delta R.T.: -0.019 min
Response: 749466
Conc: 22.71 ng/ml



#9 o-Xylene

R.T.: 13.928 min
Delta R.T.: -0.004 min
Response: 2709748
Conc: 83.97 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.208 min
Delta R.T.: -0.001 min
Response: 1973527
Conc: 73.18 ng/ml