

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB052025\
Data File : FB031790.D
Signal(s) : FID2B.CH
Acq On : 20 May 2025 15:01
Operator : YP/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: May 20 14:48:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051925.M
Quant Title :
QLast Update : Mon May 19 10:47:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 t	2-Methylpentane	33.340	31.316 E3	6.1	85	0.00
2 t	2,2,4-Trimethylpentane	39.966	39.404 E3	1.4	93	0.00
3 t	n-Heptane	38.250	37.737 E3	1.3	93	0.00
4 t	Benzene	46.891	45.642 E3	2.7	95	0.00
5 s	AAA-TFT	21.624	23.293 E3	-7.7	100	0.00
6 t	Toluene	44.153	42.809 E3	3.0	94	0.00
7 t	Ethylbenzene	38.793	37.188 E3	4.1	93	0.00
8 t	m-Xylene	43.116	41.200 E3	4.4	94	0.00
9 t	O-Xylene	39.908	38.278 E3	4.1	94	0.00
10 t	1,2,4-Trimethylbenzene	28.532	26.282 E3	7.9	92	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0