

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB052820\
 Data File : FB024404.D
 Signal(s) : FID2B.CH
 Acq On : 28 May 2020 15:39
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc : 5.04g/5.00ml DI WATER
 ALS Vial : 11 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: May 29 01:21:30 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051120.M
 Quant Title :
 QLast Update : Mon May 11 11:14:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.805	407435	18.451 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.693	779780	26.424 ng/ml
2) t 2,2,4-Trimethylpentane	7.422	839064	27.466 ng/ml
3) t n-Heptane	7.763	273218	9.019 ng/ml
4) t Benzene	7.898	373809	9.888 ng/ml
6) t Toluene	10.636	1058262	29.204 ng/ml
7) t Ethylbenzene	13.075	323253	9.535 ng/ml
8) t m-Xylene	13.208	631619	19.140 ng/ml
9) t O-Xylene	13.936	629795	19.516 ng/ml
10) t 1,2,4-Trimethylbenzene	16.213	535664	19.862 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB052820\
Data File : FB024404.D
Signal(s) : FID2B.CH
Acq On : 28 May 2020 15:39
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc : 5.04g/5.00ml DI WATER
ALS Vial : 11 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: May 29 01:21:30 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051120.M
Quant Title :
QLast Update : Mon May 11 11:14:42 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.693 min
Delta R.T.: 0.004 min
Response: 779780
Conc: 26.42 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.422 min
Delta R.T.: 0.004 min
Response: 839064
Conc: 27.47 ng/ml




#3 n-Heptane

R.T.: 7.763 min
Delta R.T.: 0.005 min
Response: 273218
Conc: 9.02 ng/ml




#4 Benzene

R.T.: 7.898 min
Delta R.T.: 0.006 min
Response: 373809
Conc: 9.89 ng/ml




#5 AAA-TFT

R.T.: 8.805 min
Delta R.T.: 0.005 min
Response: 407435
Conc: 18.45 ng/ml




#6 Toluene

R.T.: 10.636 min
Delta R.T.: 0.005 min
Response: 1058262
Conc: 29.20 ng/ml




#7 Ethylbenzene

R.T.: 13.075 min
Delta R.T.: 0.004 min
Response: 323253
Conc: 9.53 ng/ml




#8 m-Xylene

R.T.: 13.208 min
Delta R.T.: 0.004 min
Response: 631619
Conc: 19.14 ng/ml



#9 O-Xylene

R.T.: 13.936 min
Delta R.T.: 0.004 min
Response: 629795
Conc: 19.52 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.213 min
Delta R.T.: 0.003 min
Response: 535664
Conc: 19.86 ng/ml