

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB052820\
 Data File : FB024406.D
 Signal(s) : FID2B.CH
 Acq On : 28 May 2020 16:41
 Operator : DD/AJ
 Sample : L2790-10
 Misc : 5.00g/5.00ml DI WATER
 ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: May 29 01:21:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051120.M
 Quant Title :
 QLast Update : Mon May 11 11:14:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.806	489689	22.176 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.689	1779	0.060 ng/ml
3) t n-Heptane	7.750	704	0.023 ng/ml
6) t Toluene	10.638	2684	0.074 ng/ml
9) t O-Xylene	13.891	756	0.023 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB052820\
Data File : FB024406.D
Signal(s) : FID2B.CH
Acq On : 28 May 2020 16:41
Operator : DD/AJ
Sample : L2790-10
Misc : 5.00g/5.00ml DI WATER
ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: May 29 01:21:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051120.M
Quant Title :
QLast Update : Mon May 11 11:14:42 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.689 min
Delta R.T.: 0.000 min
Response: 1779
Conc: 0.06 ng/ml




#3 n-Heptane

R.T.: 7.750 min
Delta R.T.: -0.007 min
Response: 704
Conc: 0.02 ng/ml




#5 AAA-TFT

R.T.: 8.806 min
Delta R.T.: 0.006 min
Response: 489689
Conc: 22.18 ng/ml



#6 Toluene

R.T.: 10.638 min
Delta R.T.: 0.007 min
Response: 2684
Conc: 0.07 ng/ml



#9 O-Xylene

R.T.: 13.891 min
Delta R.T.: -0.041 min
Response: 756
Conc: 0.02 ng/ml