

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB052820\
 Data File : FB024399.D
 Signal(s) : FID2B.CH
 Acq On : 28 May 2020 13:06
 Operator : DD/AJ
 Sample : L2505-07MSD
 Misc : 5.05g/5.00ml DI WATER
 ALS Vial : 6 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: May 29 01:20:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051120.M
 Quant Title :
 QLast Update : Mon May 11 11:14:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.803	397853	18.017 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.692	719101	24.368 ng/ml
2) t 2,2,4-Trimethylpentane	7.421	664816	21.762 ng/ml
3) t n-Heptane	7.762	174966	5.775 ng/ml
4) t Benzene	7.897	329374	8.712 ng/ml
6) t Toluene	10.634	924290	25.507 ng/ml
7) t Ethylbenzene	13.074	278642	8.219 ng/ml
8) t m-Xylene	13.207	453328	13.737 ng/ml
9) t O-Xylene	13.934	501198	15.531 ng/ml
10) t 1,2,4-Trimethylbenzene	16.211	283622	10.516 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB052820\
Data File : FB024399.D
Signal(s) : FID2B.CH
Acq On : 28 May 2020 13:06
Operator : DD/AJ
Sample : L2505-07MSD
Misc : 5.05g/5.00ml DI WATER
ALS Vial : 6 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: May 29 01:20:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051120.M
Quant Title :
QLast Update : Mon May 11 11:14:42 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.692 min
Delta R.T.: 0.003 min
Response: 719101
Conc: 24.37 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.421 min
Delta R.T.: 0.004 min
Response: 664816
Conc: 21.76 ng/ml




#3 n-Heptane

R.T.: 7.762 min
Delta R.T.: 0.004 min
Response: 174966
Conc: 5.78 ng/ml




#4 Benzene

R.T.: 7.897 min
Delta R.T.: 0.004 min
Response: 329374
Conc: 8.71 ng/ml




#5 AAA-TFT

R.T.: 8.803 min
Delta R.T.: 0.004 min
Response: 397853
Conc: 18.02 ng/ml




#6 Toluene

R.T.: 10.634 min
Delta R.T.: 0.003 min
Response: 924290
Conc: 25.51 ng/ml




#7 Ethylbenzene

R.T.: 13.074 min
Delta R.T.: 0.003 min
Response: 278642
Conc: 8.22 ng/ml




#8 m-Xylene

R.T.: 13.207 min
Delta R.T.: 0.003 min
Response: 453328
Conc: 13.74 ng/ml



#9 O-Xylene

R.T.: 13.934 min
Delta R.T.: 0.002 min
Response: 501198
Conc: 15.53 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.211 min
Delta R.T.: 0.002 min
Response: 283622
Conc: 10.52 ng/ml