

Data Path : P:\HPCHEM1\FID_B\Data\FB052918\
 Data File : FB015815.D
 Signal(s) : FID2B.CH
 Acq On : 29 May 2018 15:16
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: May 30 04:19:03 2018
 Quant Method : P:\HPCHEM1\FID_B\Method\FB052418.M
 Quant Title :
 QLast Update : Thu May 24 11:17:17 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.118	187040	21.576 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.967	502922	33.879 ng/ml
2) t 2,2,4-Trimethylpentane	7.735	495355	33.471 ng/ml
3) t n-Heptane	8.074	158377	11.091 ng/ml
4) t Benzene	8.214	191532	10.549 ng/ml
6) t Toluene	10.967	468830	31.076 ng/ml
7) t Ethylbenzene	13.416	160380	9.727 ng/ml
8) t m-Xylene	13.549	346699	19.351 ng/ml
9) t O-Xylene	14.280	365995	19.461 ng/ml
10) t 1,2,4-Trimethylbenzene	16.561	421614	19.915 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : P:\HPCHEM1\FID_B\Data\FB052918\
Data File : FB015815.D
Signal(s) : FID2B.CH
Acq On : 29 May 2018 15:16
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: May 30 04:19:03 2018
Quant Method : P:\HPCHEM1\FID_B\Method\FB052418.M
Quant Title :
QLast Update : Thu May 24 11:17:17 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.967 min
Delta R.T.: 0.003 min
Response: 502922
Conc: 33.88 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.735 min
Delta R.T.: 0.005 min
Response: 495355
Conc: 33.47 ng/ml




#3 n-Heptane

R.T.: 8.074 min
Delta R.T.: 0.006 min
Response: 158377
Conc: 11.09 ng/ml



#4 Benzene


R.T.: 8.214 min
Delta R.T.: 0.006 min
Response: 191532
Conc: 10.55 ng/ml



#5 AAA-TFT


R.T.: 9.118 min
Delta R.T.: 0.006 min
Response: 187040
Conc: 21.58 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.967 min
Delta R.T.: 0.005 min
Response: 468830
Conc: 31.08 ng/ml




#7 Ethylbenzene

R.T.: 13.416 min
Delta R.T.: 0.005 min
Response: 160380
Conc: 9.73 ng/ml



#8 m-Xylene


R.T.: 13.549 min
Delta R.T.: 0.005 min
Response: 346699
Conc: 19.35 ng/ml



#9 O-Xylene

R.T.: 14.280 min
Delta R.T.: 0.005 min
Response: 365995
Conc: 19.46 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.561 min
Delta R.T.: 0.004 min
Response: 421614
Conc: 19.91 ng/ml