

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB053019\
 Data File : FB021256.D
 Signal(s) : FID2B.CH
 Acq On : 30 May 2019 19:17
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: May 31 01:03:21 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB052919.M
 Quant Title :
 QLast Update : Wed May 29 12:01:49 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.996	305940	20.750 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.856	1199559	31.909 ng/ml
2) t 2,2,4-Trimethylpentane	7.611	1079152	31.600 ng/ml
3) t n-Heptane	7.952	343535	10.538 ng/ml
4) t Benzene	8.088	380170	11.014 ng/ml
6) t Toluene	10.841	754789	31.845 ng/ml
7) t Ethylbenzene	13.291	243441	10.418 ng/ml
8) t m-Xylene	13.424	522129	20.919 ng/ml
9) t O-Xylene	14.155	540309	20.712 ng/ml
10) t 1,2,4-Trimethylbenzene	16.439	548976	20.442 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB053019\
Data File : FB021256.D
Signal(s) : FID2B.CH
Acq On : 30 May 2019 19:17
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 22 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: May 31 01:03:21 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB052919.M
Quant Title :
QLast Update : Wed May 29 12:01:49 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.856 min
Delta R.T.: -0.002 min
Response: 1199559
Conc: 31.91 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.611 min
Delta R.T.: -0.002 min
Response: 1079152
Conc: 31.60 ng/ml




#3 n-Heptane

R.T.: 7.952 min
Delta R.T.: -0.002 min
Response: 343535
Conc: 10.54 ng/ml




#4 Benzene

R.T.: 8.088 min
Delta R.T.: -0.002 min
Response: 380170
Conc: 11.01 ng/ml




#5 AAA-TFT

R.T.: 8.996 min
Delta R.T.: -0.002 min
Response: 305940
Conc: 20.75 ng/ml




#6 Toluene

R.T.: 10.841 min
Delta R.T.: -0.002 min
Response: 754789
Conc: 31.84 ng/ml




#7 Ethylbenzene

R.T.: 13.291 min
Delta R.T.: -0.002 min
Response: 243441
Conc: 10.42 ng/ml




#8 m-Xylene

R.T.: 13.424 min
Delta R.T.: -0.002 min
Response: 522129
Conc: 20.92 ng/ml



#9 O-Xylene

R.T.: 14.155 min
Delta R.T.: -0.002 min
Response: 540309
Conc: 20.71 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.439 min
Delta R.T.: -0.001 min
Response: 548976
Conc: 20.44 ng/ml