

Data Path : Y:\PESTPCBSRV\HPCHEM1\FID_B\Data\FB060118\
 Data File : FB015889.D
 Signal(s) : FID2B.CH
 Acq On : 1 Jun 2018 12:51
 Operator : MJ/SJ
 Sample : 100 GRO STD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jun 01 13:24:45 2018
 Quant Method : P:\HPCHEM1\FID_B\Method\FB060118.M
 Quant Title :
 QLast Update : Fri Jun 01 12:47:44 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.115	716219	102.403 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.963	1822686	150.388 ng/ml
2) t 2,2,4-Trimethylpentane	7.732	1804727	157.365 ng/ml
3) t n-Heptane	8.071	562516	51.760 ng/ml
4) t Benzene	8.210	752383	50.843 ng/ml
6) t Toluene	10.966	1863528	150.165 ng/ml
7) t Ethylbenzene	13.415	675663	47.856 ng/ml
8) t m-Xylene	13.548	1457568	94.757 ng/ml
9) t O-Xylene	14.279	1530771	97.627 ng/ml
10) t 1,2,4-Trimethylbenzene	16.561	1633819	93.661 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Y:\PESTPCBSRV\HPCHEM1\FID_B\Data\FB060118\
Data File : FB015889.D
Signal(s) : FID2B.CH
Acq On : 1 Jun 2018 12:51
Operator : MJ/SJ
Sample : 100 GRO STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jun 01 13:24:45 2018
Quant Method : P:\HPCHEM1\FID_B\Method\FB060118.M
Quant Title :
QLast Update : Fri Jun 01 12:47:44 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.963 min
Delta R.T.: -0.006 min
Response: 1822686
Conc: 150.39 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.732 min
Delta R.T.: -0.002 min
Response: 1804727
Conc: 157.36 ng/ml




#3 n-Heptane

R.T.: 8.071 min
Delta R.T.: -0.002 min
Response: 562516
Conc: 51.76 ng/ml




#4 Benzene

R.T.: 8.210 min
Delta R.T.: -0.002 min
Response: 752383
Conc: 50.84 ng/ml




#5 AAA-TFT

R.T.: 9.115 min
Delta R.T.: -0.002 min
Response: 716219
Conc: 102.40 ng/ml




#6 Toluene

R.T.: 10.966 min
Delta R.T.: 0.000 min
Response: 1863528
Conc: 150.16 ng/ml




#7 Ethylbenzene

R.T.: 13.415 min
Delta R.T.: 0.000 min
Response: 675663
Conc: 47.86 ng/ml



#8 m-Xylene

R.T.: 13.548 min
Delta R.T.: 0.000 min
Response: 1457568
Conc: 94.76 ng/ml




#9 O-Xylene

R.T.: 14.279 min

Delta R.T.: 0.000 min

Response: 1530771

Conc: 97.63 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.561 min

Delta R.T.: 0.001 min

Response: 1633819

Conc: 93.66 ng/ml