

Data Path : Y:\PESTPCBSRV\HPCHEM1\FID_B\Data\FB060118\
 Data File : FB015893.D
 Signal(s) : FID2B.CH
 Acq On : 1 Jun 2018 15:20
 Operator : MJ/SJ
 Sample : BSF0601S1
 Misc : 5.00g/5.00mL DI Water
 ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jun 02 04:02:01 2018
 Quant Method : P:\HPCHEM1\FID_B\Method\FB060118.M
 Quant Title :
 QLast Update : Fri Jun 01 13:25:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.120	148526	21.134 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.973	379030	31.257 ng/ml
2) t	2,2,4-Trimethylpentane	7.739	350678	30.280 ng/ml
3) t	n-Heptane	8.077	110680	10.113 ng/ml
4) t	Benzene	8.215	154143	10.381 ng/ml
6) t	Toluene	10.970	375699	30.268 ng/ml
7) t	Ethylbenzene	13.419	145059	10.363 ng/ml
8) t	m-Xylene	13.551	315721	20.743 ng/ml
9) t	O-Xylene	14.282	327689	20.998 ng/ml
10) t	1,2,4-Trimethylbenzene	16.564	360532	20.933 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Y:\PESTPCBSRV\HPCHEM1\FID_B\Data\FB060118\
Data File : FB015893.D
Signal(s) : FID2B.CH
Acq On : 1 Jun 2018 15:20
Operator : MJ/SJ
Sample : BSF0601S1
Misc : 5.00g/5.00mL DI Water
ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jun 02 04:02:01 2018
Quant Method : P:\HPCHEM1\FID_B\Method\FB060118.M
Quant Title :
QLast Update : Fri Jun 01 13:25:03 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.973 min
Delta R.T.: 0.004 min
Response: 379030
Conc: 31.26 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.739 min
Delta R.T.: 0.005 min
Response: 350678
Conc: 30.28 ng/ml




#3 n-Heptane

R.T.: 8.077 min
Delta R.T.: 0.004 min
Response: 110680
Conc: 10.11 ng/ml




#4 Benzene

R.T.: 8.215 min
Delta R.T.: 0.004 min
Response: 154143
Conc: 10.38 ng/ml




#5 AAA-TFT

R.T.: 9.120 min
Delta R.T.: 0.003 min
Response: 148526
Conc: 21.13 ng/ml




#6 Toluene

R.T.: 10.970 min
Delta R.T.: 0.004 min
Response: 375699
Conc: 30.27 ng/ml




#7 Ethylbenzene

R.T.: 13.419 min
Delta R.T.: 0.004 min
Response: 145059
Conc: 10.36 ng/ml




#8 m-Xylene

R.T.: 13.551 min
Delta R.T.: 0.004 min
Response: 315721
Conc: 20.74 ng/ml



#9 O-Xylene

R.T.: 14.282 min
Delta R.T.: 0.004 min
Response: 327689
Conc: 21.00 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.564 min
Delta R.T.: 0.004 min
Response: 360532
Conc: 20.93 ng/ml