

Data Path : P:\HPCHEM1\FID_B\Data\FB060118\
 Data File : FB015921.D
 Signal(s) : FID2B.CH
 Acq On : 2 Jun 2018 9:38
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jun 04 11:51:16 2018
 Quant Method : P:\HPCHEM1\FID_B\Method\FB060118.M
 Quant Title :
 QLast Update : Fri Jun 01 13:25:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.118	144122	20.508 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.970	351688	29.002 ng/ml
2) t 2,2,4-Trimethylpentane	7.736	350791	30.290 ng/ml
3) t n-Heptane	8.074	107912	9.860 ng/ml
4) t Benzene	8.212	148591	10.007 ng/ml
6) t Toluene	10.967	385668	31.071 ng/ml
7) t Ethylbenzene	13.416	147452	10.534 ng/ml
8) t m-Xylene	13.548	318890	20.951 ng/ml
9) t O-Xylene	14.280	333662	21.381 ng/ml
10) t 1,2,4-Trimethylbenzene	16.562	350908	20.375 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : P:\HPCHEM1\FID_B\Data\FB060118\
Data File : FB015921.D
Signal(s) : FID2B.CH
Acq On : 2 Jun 2018 9:38
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 29 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jun 04 11:51:16 2018
Quant Method : P:\HPCHEM1\FID_B\Method\FB060118.M
Quant Title :
QLast Update : Fri Jun 01 13:25:03 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.970 min
Delta R.T.: 0.001 min
Response: 351688
Conc: 29.00 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.736 min
Delta R.T.: 0.002 min
Response: 350791
Conc: 30.29 ng/ml




#3 n-Heptane

R.T.: 8.074 min
Delta R.T.: 0.000 min
Response: 107912
Conc: 9.86 ng/ml




#4 Benzene

R.T.: 8.212 min
Delta R.T.: 0.000 min
Response: 148591
Conc: 10.01 ng/ml




#5 AAA-TFT

R.T.: 9.118 min
Delta R.T.: 0.001 min
Response: 144122
Conc: 20.51 ng/ml




#6 Toluene

R.T.: 10.967 min
Delta R.T.: 0.002 min
Response: 385668
Conc: 31.07 ng/ml




#7 Ethylbenzene

R.T.: 13.416 min
Delta R.T.: 0.002 min
Response: 147452
Conc: 10.53 ng/ml




#8 m-Xylene

R.T.: 13.548 min
Delta R.T.: 0.001 min
Response: 318890
Conc: 20.95 ng/ml



#9 O-Xylene

R.T.: 14.280 min
Delta R.T.: 0.002 min
Response: 333662
Conc: 21.38 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.562 min
Delta R.T.: 0.002 min
Response: 350908
Conc: 20.37 ng/ml