

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB060221\
 Data File : FB025822.D
 Signal(s) : FID2B.CH
 Acq On : 2 Jun 2021 17:57
 Operator : DD/AJ
 Sample : M2578-01MS
 Misc : 5.03g/5.00ml DI WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 WASTE-CLASSMS

Integration File: Calibration.e
 Quant Time: Jun 03 01:59:44 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060221.M
 Quant Title :
 QLast Update : Wed Jun 02 14:44:25 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

	Compound	R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.218	389492	15.378 ng/ml
Target Compounds				
1) t	2-Methylpentane	5.176	604320	21.667 ng/ml
2) t	2,2,4-Trimethylpentane	7.875	370575	15.744 ng/ml
3) t	n-Heptane	8.193	191516	5.240 ng/ml
4) t	Benzene	8.344	347962	7.999 ng/ml
6) t	Toluene	11.038	953050	23.564 ng/ml
7) t	Ethylbenzene	13.445	278379	7.200 ng/ml
8) t	m-Xylene	13.575	505845	13.294 ng/ml
9) t	O-Xylene	14.298	531181	14.858 ng/ml
10) t	1,2,4-Trimethylbenzene	16.551	278326	11.087 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB060221\
Data File : FB025822.D
Signal(s) : FID2B.CH
Acq On : 2 Jun 2021 17:57
Operator : DD/AJ
Sample : M2578-01MS
Misc : 5.03g/5.00ml DI WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
WASTE-CLASSMS

Integration File: Calibration.e
Quant Time: Jun 03 01:59:44 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060221.M
Quant Title :
QLast Update : Wed Jun 02 14:44:25 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

