

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB060323\
 Data File : FB029959.D
 Signal(s) : FID2B.CH
 Acq On : 2 Jun 2023 23:42
 Operator : YP/AJ
 Sample : 10 GRO STD
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 10 GRO STD

Integration File: Calibration.e
 Quant Time: Jun 05 04:30:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060323.M
 Quant Title :
 QLast Update : Mon Jun 05 04:26:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.793	112348	7.951 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.710	263056	14.095 ng/ml
2) t 2,2,4-Trimethylpentane	7.418	262787	14.448 ng/ml
3) t n-Heptane	7.752	87941	4.718 ng/ml
4) t Benzene	7.887	101491	4.778 ng/ml
6) t Toluene	10.620	292081	14.056 ng/ml
7) t Ethylbenzene	13.057	91612	4.669 ng/ml
8) t m-Xylene	13.189	184607	9.439 ng/ml
9) t O-Xylene	13.916	180657	9.451 ng/ml
10) t 1,2,4-Trimethylbenzene	16.191	178053	9.485 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB060323\
Data File : FB029959.D
Signal(s) : FID2B.CH
Acq On : 2 Jun 2023 23:42
Operator : YP/AJ
Sample : 10 GRO STD
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
10 GRO STD

Integration File: Calibration.e
Quant Time: Jun 05 04:30:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060323.M
Quant Title :
QLast Update : Mon Jun 05 04:26:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.710 min
Delta R.T.: 0.002 min
Response: 263056
Conc: 14.10 ng/ml

Instrument :
FID_B
ClientSampleId :
10 GRO STD




#2 2,2,4-Trimethylpentane

R.T.: 7.418 min
Delta R.T.: 0.000 min
Response: 262787
Conc: 14.45 ng/ml




#3 n-Heptane

R.T.: 7.752 min
Delta R.T.: 0.000 min
Response: 87941
Conc: 4.72 ng/ml



#4 Benzene


R.T.: 7.887 min
Delta R.T.: -0.002 min
Response: 101491
Conc: 4.78 ng/ml



#5 AAA-TFT


R.T.: 8.793 min
Delta R.T.: 0.000 min
Response: 112348
Conc: 7.95 ng/ml

Instrument :
FID_B
ClientSampleId :
10 GRO STD




#6 Toluene

R.T.: 10.620 min
Delta R.T.: 0.000 min
Response: 292081
Conc: 14.06 ng/ml




#7 Ethylbenzene

R.T.: 13.057 min
Delta R.T.: 0.000 min
Response: 91612
Conc: 4.67 ng/ml



#8 m-Xylene


R.T.: 13.189 min
Delta R.T.: 0.000 min
Response: 184607
Conc: 9.44 ng/ml



#9 O-Xylene

R.T.: 13.916 min
Delta R.T.: 0.000 min
Response: 180657
Conc: 9.45 ng/ml

Instrument :
FID_B
ClientSampleId :
10 GRO STD



#10 1,2,4-Trimethylbenzene

R.T.: 16.191 min
Delta R.T.: 0.000 min
Response: 178053
Conc: 9.49 ng/ml