

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB060420\
 Data File : FB024427.D
 Signal(s) : FID2B.CH
 Acq On : 4 Jun 2020 12:33
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jun 05 00:40:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051120.M
 Quant Title :
 QLast Update : Mon May 11 11:14:42 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.802	477160	21.609 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.692	933423	31.631 ng/ml
2) t 2,2,4-Trimethylpentane	7.420	1013609	33.179 ng/ml
3) t n-Heptane	7.761	332107	10.962 ng/ml
4) t Benzene	7.895	419500	11.096 ng/ml
6) t Toluene	10.633	1202432	33.183 ng/ml
7) t Ethylbenzene	13.073	354098	10.445 ng/ml
8) t m-Xylene	13.206	690011	20.910 ng/ml
9) t O-Xylene	13.934	678795	21.035 ng/ml
10) t 1,2,4-Trimethylbenzene	16.211	571060	21.174 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB060420\
Data File : FB024427.D
Signal(s) : FID2B.CH
Acq On : 4 Jun 2020 12:33
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jun 05 00:40:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB051120.M
Quant Title :
QLast Update : Mon May 11 11:14:42 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.692 min
Delta R.T.: 0.003 min
Response: 933423
Conc: 31.63 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.420 min
Delta R.T.: 0.003 min
Response: 1013609
Conc: 33.18 ng/ml




#3 n-Heptane

R.T.: 7.761 min
Delta R.T.: 0.003 min
Response: 332107
Conc: 10.96 ng/ml




#4 Benzene

R.T.: 7.895 min
Delta R.T.: 0.002 min
Response: 419500
Conc: 11.10 ng/ml




#5 AAA-TFT

R.T.: 8.802 min
Delta R.T.: 0.003 min
Response: 477160
Conc: 21.61 ng/ml




#6 Toluene

R.T.: 10.633 min
Delta R.T.: 0.003 min
Response: 1202432
Conc: 33.18 ng/ml




#7 Ethylbenzene

R.T.: 13.073 min
Delta R.T.: 0.002 min
Response: 354098
Conc: 10.44 ng/ml




#8 m-Xylene

R.T.: 13.206 min
Delta R.T.: 0.002 min
Response: 690011
Conc: 20.91 ng/ml



#9 O-Xylene

R.T.: 13.934 min
Delta R.T.: 0.002 min
Response: 678795
Conc: 21.03 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.211 min
Delta R.T.: 0.002 min
Response: 571060
Conc: 21.17 ng/ml