

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060822\
 Data File : PS019661.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2022 02:09
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 09 03:45:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
 Quant Title : 8080.M
 QLast Update : Wed May 25 02:05:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.198	6.782	1767.7E6	598.0E6	736.810	771.702
Target Compounds							
1) T	Dalapon	3.014	2.525	3202.6E6	698.9E6	631.831	696.544
2) T	3,5-DICHL...	6.412	5.962	2182.0E6	876.6E6	690.375	731.267
3) T	4-Nitroph...	7.029	6.400	742.6E6	305.6E6	756.786	713.853
5) T	DICAMBA	7.387	6.946	6571.3E6	2533.2E6	691.875	741.027
6) T	MCPD	7.556	7.043	523.9E6	200.1E6	69.799	73.931
7) T	MCPA	7.705	7.256	738.8E6	277.9E6	68.381	72.332
8) T	DICHLORPROP	8.078	7.565	1736.0E6	646.1E6	677.983	731.422
9) T	2,4-D	8.307	7.858	2014.5E6	776.1E6	704.501	727.662
10) T	Pentachlo...	8.614	8.282	23957.4E6	8650.5E6	692.633	742.131
11) T	2,4,5-TP ...	9.164	8.643	11062.1E6	3737.9E6	704.552	743.550
12) T	2,4,5-T	9.449	9.021	13100.6E6	3502.2E6	696.111	744.641
13) T	2,4-DB	10.002	9.526	2966.1E6	458.6E6	733.093	742.156
14) T	DINOSEB	11.156	9.851	11510.6E6	2641.3E6	744.199	743.490
15) T	Picloram	10.981	10.832	21585.9E6	4511.5E6	727.354	701.387
16) T	DCPA	11.423	10.780	18078.6E6	4590.9E6	750.024	743.762



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060822\
Data File : PS019661.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2022 02:09
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 09 03:45:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
Quant Title : 8080.M
QLast Update : Wed May 25 02:05:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.014 min
Delta R.T.: -0.002 min
Response: 3202601822
Conc: 631.83 ng/ml

Instrument :
FID_B
ClientSampleId :




#1 Dalapon

R.T.: 2.525 min
Delta R.T.: -0.002 min
Response: 698941031
Conc: 696.54 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.412 min
Delta R.T.: -0.003 min
Response: 2182025493
Conc: 690.38 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 5.962 min
Delta R.T.: -0.004 min
Response: 876550689
Conc: 731.27 ng/ml



#3 4-Nitrophenol


R.T.: 7.029 min
Delta R.T.: -0.023 min
Response: 742647148
Conc: 756.79 ng/ml

Instrument :
FID_B
ClientSampleId :




#3 4-Nitrophenol

R.T.: 6.400 min
Delta R.T.: -0.010 min
Response: 305554778
Conc: 713.85 ng/ml




#4 2,4-DCAA

R.T.: 7.198 min
Delta R.T.: -0.004 min
Response: 1767729420
Conc: 736.81 ng/ml



#4 2,4-DCAA


R.T.: 6.782 min
Delta R.T.: -0.003 min
Response: 597969346
Conc: 771.70 ng/ml



#5 DICAMBA


R.T.: 7.387 min
Delta R.T.: -0.002 min
Response: 6571265364
Conc: 691.88 ng/ml

Instrument :
FID_B
ClientSampleId :




#5 DICAMBA

R.T.: 6.946 min
Delta R.T.: -0.002 min
Response: 2533161462
Conc: 741.03 ng/ml




#6 MCP

R.T.: 7.556 min
Delta R.T.: 0.004 min
Response: 523879850
Conc: 69.80 ug/ml



#6 MCP


R.T.: 7.043 min
Delta R.T.: 0.005 min
Response: 200133196
Conc: 73.93 ug/ml



#7 MCPA


R.T.: 7.705 min
Delta R.T.: 0.006 min
Response: 738786068
Conc: 68.38 ug/ml

Instrument :
FID_B
ClientSampleId :




#7 MCPA

R.T.: 7.256 min
Delta R.T.: 0.007 min
Response: 277939736
Conc: 72.33 ug/ml




#8 DICHLORPROP

R.T.: 8.078 min
Delta R.T.: -0.003 min
Response: 1736049903
Conc: 677.98 ng/ml



#8 DICHLORPROP


R.T.: 7.565 min
Delta R.T.: -0.004 min
Response: 646117828
Conc: 731.42 ng/ml



#9 2,4-D


R.T.: 8.307 min
Delta R.T.: -0.010 min
Response: 2014512976
Conc: 704.50 ng/ml

Instrument :
FID_B
ClientSampleId :




#9 2,4-D

R.T.: 7.858 min
Delta R.T.: -0.006 min
Response: 776144223
Conc: 727.66 ng/ml




#10 Pentachlorophenol

R.T.: 8.614 min
Delta R.T.: -0.001 min
Response: 23957399368
Conc: 692.63 ng/ml



#10 Pentachlorophenol


R.T.: 8.282 min
Delta R.T.: -0.002 min
Response: 8650462425
Conc: 742.13 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 9.164 min
Delta R.T.: -0.003 min
Response: 11062140011
Conc: 704.55 ng/ml

Instrument :
FID_B
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 8.643 min
Delta R.T.: -0.004 min
Response: 3737949320
Conc: 743.55 ng/ml




#12 2,4,5-T

R.T.: 9.449 min
Delta R.T.: -0.006 min
Response: 13100571654
Conc: 696.11 ng/ml



#12 2,4,5-T


R.T.: 9.021 min
Delta R.T.: -0.005 min
Response: 3502200287
Conc: 744.64 ng/ml



#13 2,4-DB


R.T.: 10.002 min
Delta R.T.: -0.009 min
Response: 2966108784
Conc: 733.09 ng/ml

Instrument :
FID_B
ClientSampleId :




#13 2,4-DB

R.T.: 9.526 min
Delta R.T.: -0.005 min
Response: 458626804
Conc: 742.16 ng/ml




#14 DINOSEB

R.T.: 11.156 min
Delta R.T.: 0.000 min
Response: 11510647711
Conc: 744.20 ng/ml



#14 DINOSEB


R.T.: 9.851 min
Delta R.T.: -0.002 min
Response: 2641335521
Conc: 743.49 ng/ml



#15 Picloram


R.T.: 10.981 min
Delta R.T.: -0.009 min
Response: 21585858840
Conc: 727.35 ng/ml

Instrument :
FID_B
ClientSampleId :




#15 Picloram

R.T.: 10.832 min
Delta R.T.: -0.005 min
Response: 4511504044
Conc: 701.39 ng/ml



#16 DCPA

R.T.: 11.423 min
Delta R.T.: -0.001 min
Response: 18078593209
Conc: 750.02 ng/ml



#16 DCPA

R.T.: 10.780 min
Delta R.T.: -0.001 min
Response: 4590861169
Conc: 743.76 ng/ml