

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB060920\
 Data File : FB024432.D
 Signal(s) : FID2B.CH
 Acq On : 9 Jun 2020 10:51
 Operator : DD/AJ
 Sample : 100 GRO STD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jun 09 11:16:49 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060920.M
 Quant Title :
 QLast Update : Tue Jun 09 10:37:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	2320907	94.962 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.683	4860910	152.605 ng/ml
2) t 2,2,4-Trimethylpentane	7.415	5132935	153.911 ng/ml
3) t n-Heptane	7.755	1698316	50.276 ng/ml
4) t Benzene	7.890	2012916	49.742 ng/ml
6) t Toluene	10.630	5762708	145.823 ng/ml
7) t Ethylbenzene	13.070	1798443	51.574 ng/ml
8) t m-Xylene	13.203	3423426	99.957 ng/ml
9) t O-Xylene	13.931	3276480	97.176 ng/ml
10) t 1,2,4-Trimethylbenzene	16.208	2763592	99.417 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB060920\
Data File : FB024432.D
Signal(s) : FID2B.CH
Acq On : 9 Jun 2020 10:51
Operator : DD/AJ
Sample : 100 GRO STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jun 09 11:16:49 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060920.M
Quant Title :
QLast Update : Tue Jun 09 10:37:58 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.683 min
Delta R.T.: -0.004 min
Response: 4860910
Conc: 152.61 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.415 min
Delta R.T.: -0.001 min
Response: 5132935
Conc: 153.91 ng/ml




#3 n-Heptane

R.T.: 7.755 min
Delta R.T.: 0.000 min
Response: 1698316
Conc: 50.28 ng/ml




#4 Benzene

R.T.: 7.890 min
Delta R.T.: 0.000 min
Response: 2012916
Conc: 49.74 ng/ml




#5 AAA-TFT

R.T.: 8.797 min
Delta R.T.: 0.000 min
Response: 2320907
Conc: 94.96 ng/ml




#6 Toluene

R.T.: 10.630 min
Delta R.T.: 0.001 min
Response: 5762708
Conc: 145.82 ng/ml




#7 Ethylbenzene

R.T.: 13.070 min
Delta R.T.: 0.002 min
Response: 1798443
Conc: 51.57 ng/ml



#8 m-Xylene

R.T.: 13.203 min
Delta R.T.: 0.002 min
Response: 3423426
Conc: 99.96 ng/ml




#9 O-Xylene

R.T.: 13.931 min

Delta R.T.: 0.002 min

Response: 3276480

Conc: 97.18 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.208 min

Delta R.T.: 0.001 min

Response: 2763592

Conc: 99.42 ng/ml