

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB060921\
 Data File : FB025864.D
 Signal(s) : FID2B.CH
 Acq On : 9 Jun 2021 19:18
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Jun 10 01:08:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060921.M
 Quant Title :
 QLast Update : Wed Jun 09 13:39:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.226	449335	18.593 ng/ml
Target Compounds			
1) t 2-Methylpentane	5.181	760825	28.910 ng/ml
2) t 2,2,4-Trimethylpentane	7.883	570250	28.307 ng/ml
3) t n-Heptane	8.201	352297	9.801 ng/ml
4) t Benzene	8.352	408364	9.782 ng/ml
6) t Toluene	11.046	1159825	28.185 ng/ml
7) t Ethylbenzene	13.452	350642	9.200 ng/ml
8) t m-Xylene	13.583	688900	17.524 ng/ml
9) t O-Xylene	14.305	595303	18.670 ng/ml
10) t 1,2,4-Trimethylbenzene	16.557	338846	17.215 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB060921\
Data File : FB025864.D
Signal(s) : FID2B.CH
Acq On : 9 Jun 2021 19:18
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Jun 10 01:08:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060921.M
Quant Title :
QLast Update : Wed Jun 09 13:39:17 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 5.181 min
Delta R.T.: 0.004 min
Response: 760825
Conc: 28.91 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.883 min
Delta R.T.: 0.005 min
Response: 570250
Conc: 28.31 ng/ml




#3 n-Heptane

R.T.: 8.201 min
Delta R.T.: 0.005 min
Response: 352297
Conc: 9.80 ng/ml



#4 Benzene


R.T.: 8.352 min
Delta R.T.: 0.005 min
Response: 408364
Conc: 9.78 ng/ml



#5 AAA-TFT


R.T.: 9.226 min
Delta R.T.: 0.005 min
Response: 449335
Conc: 18.59 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 11.046 min
Delta R.T.: 0.005 min
Response: 1159825
Conc: 28.18 ng/ml




#7 Ethylbenzene

R.T.: 13.452 min
Delta R.T.: 0.004 min
Response: 350642
Conc: 9.20 ng/ml



#8 m-Xylene


R.T.: 13.583 min
Delta R.T.: 0.004 min
Response: 688900
Conc: 17.52 ng/ml



#9 O-Xylene

R.T.: 14.305 min
Delta R.T.: 0.004 min
Response: 595303
Conc: 18.67 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.557 min
Delta R.T.: 0.004 min
Response: 338846
Conc: 17.22 ng/ml