

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB061021\
 Data File : FB025865.D
 Signal(s) : FID2B.CH
 Acq On : 10 Jun 2021 10:02
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 20 PPB GRO STD

Integration File: Calibration.e
 Quant Time: Jun 11 02:10:51 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060921.M
 Quant Title :
 QLast Update : Wed Jun 09 13:39:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.217	446988	18.496 ng/ml
Target Compounds			
1) t 2-Methylpentane	5.174	762385	28.969 ng/ml
2) t 2,2,4-Trimethylpentane	7.872	624958	31.023 ng/ml
3) t n-Heptane	8.192	348486	9.695 ng/ml
4) t Benzene	8.342	408295	9.781 ng/ml
6) t Toluene	11.037	1190544	28.931 ng/ml
7) t Ethylbenzene	13.444	363581	9.540 ng/ml
8) t m-Xylene	13.574	728067	18.520 ng/ml
9) t O-Xylene	14.297	628115	19.699 ng/ml
10) t 1,2,4-Trimethylbenzene	16.549	390326	19.831 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB061021\
Data File : FB025865.D
Signal(s) : FID2B.CH
Acq On : 10 Jun 2021 10:02
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD

Integration File: Calibration.e
Quant Time: Jun 11 02:10:51 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060921.M
Quant Title :
QLast Update : Wed Jun 09 13:39:17 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 5.174 min
Delta R.T.: -0.003 min
Response: 762385
Conc: 28.97 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#2 2,2,4-Trimethylpentane

R.T.: 7.872 min
Delta R.T.: -0.006 min
Response: 624958
Conc: 31.02 ng/ml




#3 n-Heptane

R.T.: 8.192 min
Delta R.T.: -0.004 min
Response: 348486
Conc: 9.69 ng/ml



#4 Benzene


R.T.: 8.342 min
Delta R.T.: -0.004 min
Response: 408295
Conc: 9.78 ng/ml



#5 AAA-TFT


R.T.: 9.217 min
Delta R.T.: -0.004 min
Response: 446988
Conc: 18.50 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#6 Toluene

R.T.: 11.037 min
Delta R.T.: -0.005 min
Response: 1190544
Conc: 28.93 ng/ml




#7 Ethylbenzene

R.T.: 13.444 min
Delta R.T.: -0.005 min
Response: 363581
Conc: 9.54 ng/ml



#8 m-Xylene


R.T.: 13.574 min
Delta R.T.: -0.005 min
Response: 728067
Conc: 18.52 ng/ml



#9 O-Xylene

R.T.: 14.297 min
Delta R.T.: -0.004 min
Response: 628115
Conc: 19.70 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD



#10 1,2,4-Trimethylbenzene

R.T.: 16.549 min
Delta R.T.: -0.004 min
Response: 390326
Conc: 19.83 ng/ml