

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB061422\
 Data File : FB027432.D
 Signal(s) : FID2B.CH
 Acq On : 14 Jun 2022 17:39
 Operator : AJ/MA
 Sample : BSF0614S2
 Misc : 5.03G/5.00 ML DI WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF0614S2

Integration File: SAMPLE.e
 Quant Time: Jun 15 03:14:13 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB061022.M
 Quant Title :
 QLast Update : Mon Jun 13 09:22:45 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.795	236900	21.120 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.711	500228	29.692 ng/ml
2) t 2,2,4-Trimethylpentane	7.427	525018	29.901 ng/ml
3) t n-Heptane	7.753	171560	9.793 ng/ml
4) t Benzene	7.890	200262	9.820 ng/ml
6) t Toluene	10.623	574159	29.654 ng/ml
7) t Ethylbenzene	13.060	178206	9.841 ng/ml
8) t m-Xylene	13.193	353222	19.815 ng/ml
9) t O-Xylene	13.920	343214	19.752 ng/ml
10) t 1,2,4-Trimethylbenzene	16.195	342929	19.963 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB061422\
Data File : FB027432.D
Signal(s) : FID2B.CH
Acq On : 14 Jun 2022 17:39
Operator : AJ/MA
Sample : BSF0614S2
Misc : 5.03G/5.00 ML DI WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF0614S2

Integration File: SAMPLE.e
Quant Time: Jun 15 03:14:13 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB061022.M
Quant Title :
QLast Update : Mon Jun 13 09:22:45 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

