

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB061720\
 Data File : FB024469.D
 Signal(s) : FID2B.CH
 Acq On : 17 Jun 2020 12:17
 Operator : DD/AJ
 Sample : L2811-02MSD
 Misc : 5.03gm/5.00ml DI WATER
 ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jun 18 03:21:52 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060920.M
 Quant Title :
 QLast Update : Tue Jun 09 11:17:18 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.796	412094	17.033 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.688	679542	21.260 ng/ml
2) t 2,2,4-Trimethylpentane	7.415	645507	19.255 ng/ml
3) t n-Heptane	7.755	213369	6.310 ng/ml
4) t Benzene	7.889	336068	8.313 ng/ml
6) t Toluene	10.627	929663	23.657 ng/ml
7) t Ethylbenzene	13.067	277778	7.916 ng/ml
8) t m-Xylene	13.199	541765	15.820 ng/ml
9) t O-Xylene	13.927	541394	16.148 ng/ml
10) t 1,2,4-Trimethylbenzene	16.204	424319	15.282 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB061720\
Data File : FB024469.D
Signal(s) : FID2B.CH
Acq On : 17 Jun 2020 12:17
Operator : DD/AJ
Sample : L2811-02MSD
Misc : 5.03gm/5.00ml DI WATER
ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jun 18 03:21:52 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060920.M
Quant Title :
QLast Update : Tue Jun 09 11:17:18 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.688 min
Delta R.T.: 0.000 min
Response: 679542
Conc: 21.26 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.415 min
Delta R.T.: 0.000 min
Response: 645507
Conc: 19.26 ng/ml




#3 n-Heptane

R.T.: 7.755 min
Delta R.T.: 0.000 min
Response: 213369
Conc: 6.31 ng/ml




#4 Benzene

R.T.: 7.889 min
Delta R.T.: 0.000 min
Response: 336068
Conc: 8.31 ng/ml




#5 AAA-TFT

R.T.: 8.796 min
Delta R.T.: -0.001 min
Response: 412094
Conc: 17.03 ng/ml




#6 Toluene

R.T.: 10.627 min
Delta R.T.: -0.001 min
Response: 929663
Conc: 23.66 ng/ml




#7 Ethylbenzene

R.T.: 13.067 min
Delta R.T.: -0.001 min
Response: 277778
Conc: 7.92 ng/ml




#8 m-Xylene

R.T.: 13.199 min
Delta R.T.: -0.002 min
Response: 541765
Conc: 15.82 ng/ml



#9 O-Xylene

R.T.: 13.927 min
Delta R.T.: -0.002 min
Response: 541394
Conc: 16.15 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.204 min
Delta R.T.: -0.003 min
Response: 424319
Conc: 15.28 ng/ml