

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB061919\
 Data File : FB021432.D
 Signal(s) : FID2B.CH
 Acq On : 19 Jun 2019 10:32
 Operator : MJ/SJ
 Sample : BSF0619S1
 Misc : 5.01g/5.00ml DI WATER
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jun 19 14:56:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB052919.M
 Quant Title :
 QLast Update : Wed May 29 12:01:49 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.992	267325	18.131 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.865	742080	19.740 ng/ml
2) t 2,2,4-Trimethylpentane	7.609	903734	26.463 ng/ml
3) t n-Heptane	7.948	257813	7.908 ng/ml
4) t Benzene	8.089	358328	10.381 ng/ml
6) t Toluene	10.833	737653	31.122 ng/ml
7) t Ethylbenzene	13.276	212298	9.085 ng/ml
8) t m-Xylene	13.408	464411	18.606 ng/ml
9) t O-Xylene	14.138	468589	17.963 ng/ml
10) t 1,2,4-Trimethylbenzene	16.417	468467	17.444 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB061919\
Data File : FB021432.D
Signal(s) : FID2B.CH
Acq On : 19 Jun 2019 10:32
Operator : MJ/SJ
Sample : BSF0619S1
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jun 19 14:56:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB052919.M
Quant Title :
QLast Update : Wed May 29 12:01:49 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

