

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB062123\  
 Data File : FB030070.D  
 Signal(s) : FID2B.CH  
 Acq On : 21 Jun 2023 13:42  
 Operator : YP/AJ  
 Sample : 03352-01MS 50X  
 Misc : 5.01G/5.00 ML DI WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 FID\_B  
 ClientSampleId :  
 BU-3-062023MS

Integration File: Calibration.e  
 Quant Time: Jun 21 23:24:55 2023  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB060323.M  
 Quant Title :  
 QLast Update : Mon Jun 05 04:33:00 2023  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 5 g/ml  
 Signal Phase : RTX-502.2  
 Signal Info : 60mx0.53mmx3.00um

| Compound                     | R.T.   | Response | Conc Units   |
|------------------------------|--------|----------|--------------|
| -----                        |        |          |              |
| System Monitoring Compounds  |        |          |              |
| 5) s AAA-TFT                 | 8.784  | 296675   | 21.324 ng/ml |
| Target Compounds             |        |          |              |
| 1) t 2-Methylpentane         | 4.696  | 507588   | 26.865 ng/ml |
| 2) t 2,2,4-Trimethylpentane  | 7.410  | 531213   | 28.152 ng/ml |
| 3) t n-Heptane               | 7.740  | 193221   | 9.982 ng/ml  |
| 4) t Benzene                 | 7.877  | 243673   | 10.886 ng/ml |
| 6) t Toluene                 | 10.612 | 698763   | 32.196 ng/ml |
| 7) t Ethylbenzene            | 13.049 | 216417   | 10.703 ng/ml |
| 8) t m-Xylene                | 13.182 | 424595   | 21.113 ng/ml |
| 9) t O-Xylene                | 13.909 | 405229   | 20.791 ng/ml |
| 10) t 1,2,4-Trimethylbenzene | 16.183 | 368087   | 19.652 ng/ml |
| -----                        |        |          |              |

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB062123\  
Data File : FB030070.D  
Signal(s) : FID2B.CH  
Acq On : 21 Jun 2023 13:42  
Operator : YP/AJ  
Sample : 03352-01MS 50X  
Misc : 5.01G/5.00 ML DI WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
FID\_B  
ClientSampleId :  
BU-3-062023MS

Integration File: Calibration.e  
Quant Time: Jun 21 23:24:55 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB060323.M  
Quant Title :  
QLast Update : Mon Jun 05 04:33:00 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 5 g/ml  
Signal Phase : RTX-502.2  
Signal Info : 60mx0.53mmx3.00um

