

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB062123\  
 Data File : FB030114.D  
 Signal(s) : FID2B.CH  
 Acq On : 22 Jun 2023 20:10  
 Operator : YP/AJ  
 Sample : 20 PPB GRO STD  
 Misc :  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 FID\_B  
 ClientSampleId :  
 20 PPB GRO STD

Integration File: Calibration.e  
 Quant Time: Jun 22 23:37:57 2023  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB060323.M  
 Quant Title :  
 QLast Update : Mon Jun 05 04:33:00 2023  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 5 g/ml  
 Signal Phase : RTX-502.2  
 Signal Info : 60mx0.53mmx3.00um

| Compound                     | R.T.   | Response | Conc Units   |
|------------------------------|--------|----------|--------------|
| -----                        |        |          |              |
| System Monitoring Compounds  |        |          |              |
| 5) s AAA-TFT                 | 8.788  | 268623   | 19.308 ng/ml |
| Target Compounds             |        |          |              |
| 1) t 2-Methylpentane         | 4.706  | 524767   | 27.774 ng/ml |
| 2) t 2,2,4-Trimethylpentane  | 7.415  | 577027   | 30.580 ng/ml |
| 3) t n-Heptane               | 7.748  | 176968   | 9.143 ng/ml  |
| 4) t Benzene                 | 7.884  | 222340   | 9.933 ng/ml  |
| 6) t Toluene                 | 10.617 | 637902   | 29.392 ng/ml |
| 7) t Ethylbenzene            | 13.053 | 207514   | 10.263 ng/ml |
| 8) t m-Xylene                | 13.186 | 410869   | 20.431 ng/ml |
| 9) t O-Xylene                | 13.912 | 400951   | 20.572 ng/ml |
| 10) t 1,2,4-Trimethylbenzene | 16.187 | 395511   | 21.116 ng/ml |
| -----                        |        |          |              |

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB062123\  
Data File : FB030114.D  
Signal(s) : FID2B.CH  
Acq On : 22 Jun 2023 20:10  
Operator : YP/AJ  
Sample : 20 PPB GRO STD  
Misc :  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
FID\_B  
ClientSampleId :  
20 PPB GRO STD

Integration File: Calibration.e  
Quant Time: Jun 22 23:37:57 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB060323.M  
Quant Title :  
QLast Update : Mon Jun 05 04:33:00 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 5 g/ml  
Signal Phase : RTX-502.2  
Signal Info : 60mx0.53mmx3.00um

