

rteres

Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB062123\
Data File : FB030072.D
Signal(s) : FID2B.CH
Acq On : 21 Jun 2023 14:59
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060323.M
Title :

Signal : FID2B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.703	4.520	4.839	PV	7540	569448	82.00%	14.219%
2	7.412	7.196	7.623	PV	6866	580254	83.56%	14.489%
3	7.747	7.623	7.813	VV	3742	204455	29.44%	5.105%
4	7.882	7.813	8.016	VV	4726	242107	34.86%	6.045%
5	8.787	8.606	8.968	BV	4730	260383	37.49%	6.502%
6	10.614	10.505	10.763	VV	15317	694453	100.00%	17.340%
7	13.051	12.795	13.112	BV	4998	218254	31.43%	5.450%
8	13.184	13.112	13.374	VV	9817	431228	62.10%	10.768%
9	13.910	13.597	14.028	BV	9271	410668	59.14%	10.254%
10	16.184	16.038	16.322	VV	9371	393640	56.68%	9.829%

Sum of corrected areas: 4004892

FB060323.M Tue Jul 04 06:40:59 2023