

Instrument :
FID_B
LabSampleId :
20 PPB GRO STD

rteres

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB
Data File : FB030101.D
Signal(s) : FID2B.CH
Acq On : 22 Jun 2023 11:47
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 43 Sample Multiplier: 1

Reviewed By :Yogesh Patel 06/23/2023
Supervised By :Ankita Jodhani 06/23/2023

Integration File: Calibration.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060323.M
Title :

Signal : FID2B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.702	4.520	4.842	PV	7817	588383	89.79%	14.751%
2	7.410	7.175	7.611	PV	6998	591667	90.29%	14.833%
3	7.745	7.611	7.812	VV	3666	200196	30.55%	5.019%
4	7.881	7.812	8.019	VV	4488	229773	35.06%	5.760%
5	8.785	8.560	8.932	PV	4899	269702	41.16%	6.761%
6	10.613	10.502	10.751	VV	14438	655291	100.00%	16.428%
7	13.049	12.790	13.110	BV	4753	211659	32.30%	5.306%
8	13.181	13.110	13.352	VV	9211	404234	61.69%	10.134%
9	13.908	13.424	14.010	VV	8747	403409	61.56%	10.113%
10	16.183	16.080	16.322	VV	9304	434574	66.32%	10.895%

Sum of corrected areas: 3988886

FB060323.M Thu Jun 22 23:54:30 2023