

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB062725\
 Data File : FB031948.D
 Signal(s) : FID2B.CH
 Acq On : 27 Jun 2025 11:19
 Operator : YP/AJ
 Sample : BSF0627S1
 Misc : 5.00G/5.00 ML DI WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF0627S1

Integration File: Calibration.e
 Quant Time: Jun 28 05:14:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB062325.M
 Quant Title :
 QLast Update : Mon Jun 23 13:56:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

	Compound	R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	8.795	258901	16.781 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.708	610924	24.014 ng/ml
2) t	2,2,4-Trimethylpentane	7.420	759296	28.156 ng/ml
3) t	n-Heptane	7.754	223235	8.947 ng/ml
4) t	Benzene	7.892	302492	8.897 ng/ml
6) t	Toluene	10.623	903191	25.259 ng/ml
7) t	Ethylbenzene	13.060	264115	8.364 ng/ml
8) t	m-Xylene	13.193	551132	16.803 ng/ml
9) t	O-Xylene	13.920	522882	16.847 ng/ml
10) t	1,2,4-Trimethylbenzene	16.195	438278	16.576 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB062725\
Data File : FB031948.D
Signal(s) : FID2B.CH
Acq On : 27 Jun 2025 11:19
Operator : YP/AJ
Sample : BSF0627S1
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF0627S1

Integration File: Calibration.e
Quant Time: Jun 28 05:14:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB062325.M
Quant Title :
QLast Update : Mon Jun 23 13:56:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

