

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB062818\
 Data File : FB016466.D
 Signal(s) : FID2B.CH
 Acq On : 28 Jun 2018 21:24
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Jun 29 13:54:44 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060118.M
 Quant Title :
 QLast Update : Thu Jun 28 12:22:11 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.119	115764	16.473 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.972	370531	30.556 ng/ml
2) t 2,2,4-Trimethylpentane	7.737	349393	30.169 ng/ml
3) t n-Heptane	8.073	113766	10.395 ng/ml
4) t Benzene	8.214	148339	9.990 ng/ml
6) t Toluene	10.968	325967	26.261 ng/ml
7) t Ethylbenzene	13.417	120783	8.629 ng/ml
8) t m-Xylene	13.550	259289	17.035 ng/ml
9) t O-Xylene	14.282	285366	18.286 ng/ml
10) t 1,2,4-Trimethylbenzene	16.564	300518	17.449 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB062818\
Data File : FB016466.D
Signal(s) : FID2B.CH
Acq On : 28 Jun 2018 21:24
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Jun 29 13:54:44 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060118.M
Quant Title :
QLast Update : Thu Jun 28 12:22:11 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.972 min
Delta R.T.: 0.003 min
Response: 370531
Conc: 30.56 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.737 min
Delta R.T.: 0.003 min
Response: 349393
Conc: 30.17 ng/ml




#3 n-Heptane

R.T.: 8.073 min
Delta R.T.: 0.000 min
Response: 113766
Conc: 10.39 ng/ml



#4 Benzene


R.T.: 8.214 min
Delta R.T.: 0.003 min
Response: 148339
Conc: 9.99 ng/ml



#5 AAA-TFT


R.T.: 9.119 min
Delta R.T.: 0.003 min
Response: 115764
Conc: 16.47 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.968 min
Delta R.T.: 0.003 min
Response: 325967
Conc: 26.26 ng/ml




#7 Ethylbenzene

R.T.: 13.417 min
Delta R.T.: 0.002 min
Response: 120783
Conc: 8.63 ng/ml



#8 m-Xylene


R.T.: 13.550 min
Delta R.T.: 0.003 min
Response: 259289
Conc: 17.04 ng/ml



#9 O-Xylene

R.T.: 14.282 min
Delta R.T.: 0.003 min
Response: 285366
Conc: 18.29 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.564 min
Delta R.T.: 0.003 min
Response: 300518
Conc: 17.45 ng/ml