

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070121\
 Data File : FB025973.D
 Signal(s) : FID2B.CH
 Acq On : 1 Jul 2021 12:08
 Operator : AJ/MA
 Sample : M2857-02
 Misc : 5.08g/5.00 ml DI WATER
 ALS Vial : 4 Sample Multiplier: 1

Integration File: SAMPLE.e
 Quant Time: Jul 01 13:55:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060921.M
 Quant Title :
 QLast Update : Wed Jun 09 13:39:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

	Compound	R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.223	465307	19.254 ng/ml
Target Compounds				
1) t	2-Methylpentane	5.168	950	0.036 ng/ml
2) t	2,2,4-Trimethylpentane	7.867	8463	0.420 ng/ml
3) t	n-Heptane	8.201	1118	0.031 ng/ml
4) t	Benzene	8.365	1727	0.041 ng/ml
6) t	Toluene	11.042	6924	0.168 ng/ml
7) t	Ethylbenzene	13.445	2322	0.061 ng/ml
8) t	m-Xylene	13.576	4157	0.106 ng/ml
9) t	O-Xylene	14.307	3133	0.098 ng/ml
10) t	1,2,4-Trimethylbenzene	16.559	4908	0.249 ng/ml


(f)=RT Delta > 1/2 window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070121\
Data File : FB025973.D
Signal(s) : FID2B.CH
Acq On : 1 Jul 2021 12:08
Operator : AJ/MA
Sample : M2857-02
Misc : 5.08g/5.00 ml DI WATER
ALS Vial : 4 Sample Multiplier: 1

Integration File: SAMPLE.e
Quant Time: Jul 01 13:55:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060921.M
Quant Title :
QLast Update : Wed Jun 09 13:39:17 2021
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 5.168 min
Delta R.T.: -0.009 min
Response: 950
Conc: 0.04 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.867 min
Delta R.T.: -0.011 min
Response: 8463
Conc: 0.42 ng/ml




#3 n-Heptane

R.T.: 8.201 min
Delta R.T.: 0.005 min
Response: 1118
Conc: 0.03 ng/ml




#4 Benzene

R.T.: 8.365 min
Delta R.T.: 0.018 min
Response: 1727
Conc: 0.04 ng/ml




#5 AAA-TFT

R.T.: 9.223 min
Delta R.T.: 0.002 min
Response: 465307
Conc: 19.25 ng/ml




#6 Toluene

R.T.: 11.042 min
Delta R.T.: 0.001 min
Response: 6924
Conc: 0.17 ng/ml




#7 Ethylbenzene

R.T.: 13.445 min
Delta R.T.: -0.003 min
Response: 2322
Conc: 0.06 ng/ml




#8 m-Xylene

R.T.: 13.576 min
Delta R.T.: -0.002 min
Response: 4157
Conc: 0.11 ng/ml



#9 O-Xylene

R.T.: 14.307 min
Delta R.T.: 0.006 min
Response: 3133
Conc: 0.10 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.559 min
Delta R.T.: 0.005 min
Response: 4908
Conc: 0.25 ng/ml