

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070122\
 Data File : FB027599.D
 Signal(s) : FID2B.CH
 Acq On : 1 Jul 2022 17:54
 Operator : AJ/MA
 Sample : N3564-14 50X
 Misc : 5.09G/5.00 ML MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Jul 02 05:05:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB061022.M
 Quant Title :
 QLast Update : Mon Jun 13 09:22:45 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

Target Compounds				
1) t 2-Methylpentane	4.711	2053390	121.883	ng/ml
2) t 2,2,4-Trimethylpentane	7.386	2989921	170.285	ng/ml
3) t n-Heptane	7.760	6757590	385.734	ng/ml
7) t Ethylbenzene	13.020	4453853	245.941	ng/ml
8) t m-Xylene	13.188	532517	29.873	ng/ml
9) t O-Xylene	13.930	376399	21.662	ng/ml
10) t 1,2,4-Trimethylbenzene	16.205	253717	14.770	ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070122\
Data File : FB027599.D
Signal(s) : FID2B.CH
Acq On : 1 Jul 2022 17:54
Operator : AJ/MA
Sample : N3564-14 50X
Misc : 5.09G/5.00 ML MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Jul 02 05:05:42 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB061022.M
Quant Title :
QLast Update : Mon Jun 13 09:22:45 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.711 min
Delta R.T.: 0.007 min
Response: 2053390
Conc: 121.88 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.386 min
Delta R.T.: -0.040 min
Response: 2989921
Conc: 170.28 ng/ml




#3 n-Heptane

R.T.: 7.760 min
Delta R.T.: 0.011 min
Response: 6757590
Conc: 385.73 ng/ml



#5 AAA-TFT


R.T.: 0.000 min
Exp R.T. : 8.791 min
Response: 0
Conc: N.D.



#7 Ethylbenzene


R.T.: 13.020 min
Delta R.T.: -0.036 min
Response: 4453853
Conc: 245.94 ng/ml

Instrument :
FID_B
ClientSampleId :




#8 m-Xylene

R.T.: 13.188 min
Delta R.T.: -0.001 min
Response: 532517
Conc: 29.87 ng/ml



#9 o-Xylene

R.T.: 13.930 min
Delta R.T.: 0.014 min
Response: 376399
Conc: 21.66 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.205 min
Delta R.T.: 0.013 min
Response: 253717
Conc: 14.77 ng/ml