

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070122\
 Data File : FB027591.D
 Signal(s) : FID2B.CH
 Acq On : 1 Jul 2022 12:48
 Operator : AJ/MA
 Sample : BSF0701S2
 Misc : 5.00G/5.00 ML DI WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF0701S2

Integration File: Calibration.e
 Quant Time: Jul 02 05:04:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB061022.M
 Quant Title :
 QLast Update : Mon Jun 13 09:22:45 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.796	253261	22.579 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.714	483704	28.711 ng/ml
2) t 2,2,4-Trimethylpentane	7.430	513657	29.254 ng/ml
3) t n-Heptane	7.754	161466	9.217 ng/ml
4) t Benzene	7.890	191714	9.401 ng/ml
6) t Toluene	10.624	535278	27.646 ng/ml
7) t Ethylbenzene	13.061	165571	9.143 ng/ml
8) t m-Xylene	13.193	332089	18.629 ng/ml
9) t O-Xylene	13.920	327400	18.842 ng/ml
10) t 1,2,4-Trimethylbenzene	16.194	327248	19.050 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070122\
Data File : FB027591.D
Signal(s) : FID2B.CH
Acq On : 1 Jul 2022 12:48
Operator : AJ/MA
Sample : BSF0701S2
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF0701S2

Integration File: Calibration.e
Quant Time: Jul 02 05:04:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB061022.M
Quant Title :
QLast Update : Mon Jun 13 09:22:45 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

