

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070219\
 Data File : FB021625.D
 Signal(s) : FID2B.CH
 Acq On : 2 Jul 2019 16:22
 Operator : MJ/SJ
 Sample : BSF0702W2
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jul 03 01:30:29 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB062419.M
 Quant Title :
 QLast Update : Mon Jun 24 16:18:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

	Compound	R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	8.990	273710	19.796 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.861	1000969	29.322 ng/ml
2) t	2,2,4-Trimethylpentane	7.610	955980	30.935 ng/ml
3) t	n-Heptane	7.949	298938	10.214 ng/ml
4) t	Benzene	8.084	349460	10.412 ng/ml
6) t	Toluene	10.831	700759	30.678 ng/ml
7) t	Ethylbenzene	13.277	235791	9.960 ng/ml
8) t	m-Xylene	13.410	504604	19.909 ng/ml
9) t	O-Xylene	14.139	524178	19.986 ng/ml
10) t	1,2,4-Trimethylbenzene	16.419	534244	20.334 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070219\
Data File : FB021625.D
Signal(s) : FID2B.CH
Acq On : 2 Jul 2019 16:22
Operator : MJ/SJ
Sample : BSF0702W2
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jul 03 01:30:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB062419.M
Quant Title :
QLast Update : Mon Jun 24 16:18:04 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.861 min
Delta R.T.: 0.000 min
Response: 1000969
Conc: 29.32 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.610 min
Delta R.T.: 0.002 min
Response: 955980
Conc: 30.93 ng/ml




#3 n-Heptane

R.T.: 7.949 min
Delta R.T.: 0.000 min
Response: 298938
Conc: 10.21 ng/ml




#4 Benzene

R.T.: 8.084 min
Delta R.T.: 0.002 min
Response: 349460
Conc: 10.41 ng/ml




#5 AAA-TFT

R.T.: 8.990 min
Delta R.T.: 0.002 min
Response: 273710
Conc: 19.80 ng/ml




#6 Toluene

R.T.: 10.831 min
Delta R.T.: 0.002 min
Response: 700759
Conc: 30.68 ng/ml




#7 Ethylbenzene

R.T.: 13.277 min
Delta R.T.: 0.001 min
Response: 235791
Conc: 9.96 ng/ml




#8 m-Xylene

R.T.: 13.410 min
Delta R.T.: 0.002 min
Response: 504604
Conc: 19.91 ng/ml



#9 O-Xylene

R.T.: 14.139 min
Delta R.T.: 0.002 min
Response: 524178
Conc: 19.99 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.419 min
Delta R.T.: 0.002 min
Response: 534244
Conc: 20.33 ng/ml