

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070319\
 Data File : FB021629.D
 Signal(s) : FID2B.CH
 Acq On : 3 Jul 2019 12:30
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jul 03 22:25:34 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB062419.M
 Quant Title :
 QLast Update : Mon Jun 24 16:18:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	8.985	245444	17.752 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.858	854859	25.042 ng/ml
2) t	2,2,4-Trimethylpentane	7.604	800820	25.914 ng/ml
3) t	n-Heptane	7.944	251067	8.578 ng/ml
4) t	Benzene	8.079	294741	8.781 ng/ml
6) t	Toluene	10.826	633578	27.737 ng/ml
7) t	Ethylbenzene	13.272	232268	9.811 ng/ml
8) t	m-Xylene	13.405	499110	19.692 ng/ml
9) t	O-Xylene	14.134	519836	19.821 ng/ml
10) t	1,2,4-Trimethylbenzene	16.414	532850	20.281 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070319\
Data File : FB021629.D
Signal(s) : FID2B.CH
Acq On : 3 Jul 2019 12:30
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jul 03 22:25:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB062419.M
Quant Title :
QLast Update : Mon Jun 24 16:18:04 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.858 min
Delta R.T.: -0.003 min
Response: 854859
Conc: 25.04 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.604 min
Delta R.T.: -0.004 min
Response: 800820
Conc: 25.91 ng/ml




#3 n-Heptane

R.T.: 7.944 min
Delta R.T.: -0.004 min
Response: 251067
Conc: 8.58 ng/ml




#4 Benzene

R.T.: 8.079 min
Delta R.T.: -0.003 min
Response: 294741
Conc: 8.78 ng/ml




#5 AAA-TFT

R.T.: 8.985 min
Delta R.T.: -0.004 min
Response: 245444
Conc: 17.75 ng/ml




#6 Toluene

R.T.: 10.826 min
Delta R.T.: -0.004 min
Response: 633578
Conc: 27.74 ng/ml




#7 Ethylbenzene

R.T.: 13.272 min
Delta R.T.: -0.003 min
Response: 232268
Conc: 9.81 ng/ml




#8 m-Xylene

R.T.: 13.405 min
Delta R.T.: -0.003 min
Response: 499110
Conc: 19.69 ng/ml



#9 O-Xylene

R.T.: 14.134 min
Delta R.T.: -0.003 min
Response: 519836
Conc: 19.82 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.414 min
Delta R.T.: -0.003 min
Response: 532850
Conc: 20.28 ng/ml