

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070920\
 Data File : FB024558.D
 Signal(s) : FID2B.CH
 Acq On : 9 Jul 2020 10:59
 Operator : DD/AJ
 Sample : FB070920GROICV
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Jul 09 11:22:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070920.M
 Quant Title :
 QLast Update : Thu Jul 09 10:49:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.799	457256	19.850 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.688	893792	30.203 ng/ml
2) t 2,2,4-Trimethylpentane	7.418	919109	29.358 ng/ml
3) t n-Heptane	7.759	305316	9.820 ng/ml
4) t Benzene	7.893	387801	10.174 ng/ml
6) t Toluene	10.631	1090134	29.814 ng/ml
7) t Ethylbenzene	13.070	330170	9.666 ng/ml
8) t m-Xylene	13.203	653819	19.374 ng/ml
9) t O-Xylene	13.931	644174	19.599 ng/ml
10) t 1,2,4-Trimethylbenzene	16.209	541049	20.058 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070920\
Data File : FB024558.D
Signal(s) : FID2B.CH
Acq On : 9 Jul 2020 10:59
Operator : DD/AJ
Sample : FB070920GROICV
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Jul 09 11:22:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070920.M
Quant Title :
QLast Update : Thu Jul 09 10:49:06 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.688 min
Delta R.T.: 0.000 min
Response: 893792
Conc: 30.20 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.418 min
Delta R.T.: 0.002 min
Response: 919109
Conc: 29.36 ng/ml




#3 n-Heptane

R.T.: 7.759 min
Delta R.T.: 0.003 min
Response: 305316
Conc: 9.82 ng/ml



#4 Benzene


R.T.: 7.893 min
Delta R.T.: 0.000 min
Response: 387801
Conc: 10.17 ng/ml



#5 AAA-TFT


R.T.: 8.799 min
Delta R.T.: 0.001 min
Response: 457256
Conc: 19.85 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.631 min
Delta R.T.: 0.002 min
Response: 1090134
Conc: 29.81 ng/ml




#7 Ethylbenzene

R.T.: 13.070 min
Delta R.T.: 0.001 min
Response: 330170
Conc: 9.67 ng/ml



#8 m-Xylene


R.T.: 13.203 min
Delta R.T.: 0.002 min
Response: 653819
Conc: 19.37 ng/ml



#9 O-Xylene

R.T.: 13.931 min
Delta R.T.: 0.002 min
Response: 644174
Conc: 19.60 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.209 min
Delta R.T.: 0.002 min
Response: 541049
Conc: 20.06 ng/ml