

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB071019\
 Data File : FB021657.D
 Signal(s) : FID2B.CH
 Acq On : 10 Jul 2019 8:46
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jul 11 01:45:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB062419.M
 Quant Title :
 QLast Update : Mon Jun 24 16:18:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	8.990	264295	19.115 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.861	875754	25.654 ng/ml
2) t	2,2,4-Trimethylpentane	7.610	816829	26.432 ng/ml
3) t	n-Heptane	7.949	258554	8.834 ng/ml
4) t	Benzene	8.084	316740	9.437 ng/ml
6) t	Toluene	10.831	666580	29.182 ng/ml
7) t	Ethylbenzene	13.277	240503	10.159 ng/ml
8) t	m-Xylene	13.410	516837	20.391 ng/ml
9) t	O-Xylene	14.139	537381	20.490 ng/ml
10) t	1,2,4-Trimethylbenzene	16.419	548473	20.875 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB071019\
Data File : FB021657.D
Signal(s) : FID2B.CH
Acq On : 10 Jul 2019 8:46
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jul 11 01:45:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB062419.M
Quant Title :
QLast Update : Mon Jun 24 16:18:04 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.861 min
Delta R.T.: 0.000 min
Response: 875754
Conc: 25.65 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.610 min
Delta R.T.: 0.002 min
Response: 816829
Conc: 26.43 ng/ml




#3 n-Heptane

R.T.: 7.949 min
Delta R.T.: 0.001 min
Response: 258554
Conc: 8.83 ng/ml




#4 Benzene

R.T.: 8.084 min
Delta R.T.: 0.002 min
Response: 316740
Conc: 9.44 ng/ml




#5 AAA-TFT

R.T.: 8.990 min
Delta R.T.: 0.002 min
Response: 264295
Conc: 19.12 ng/ml




#6 Toluene

R.T.: 10.831 min
Delta R.T.: 0.002 min
Response: 666580
Conc: 29.18 ng/ml




#7 Ethylbenzene

R.T.: 13.277 min
Delta R.T.: 0.002 min
Response: 240503
Conc: 10.16 ng/ml




#8 m-Xylene

R.T.: 13.410 min
Delta R.T.: 0.002 min
Response: 516837
Conc: 20.39 ng/ml



#9 O-Xylene

R.T.: 14.139 min
Delta R.T.: 0.002 min
Response: 537381
Conc: 20.49 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.419 min
Delta R.T.: 0.002 min
Response: 548473
Conc: 20.88 ng/ml