

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB071123\
 Data File : FB030152.D
 Signal(s) : FID2B.CH
 Acq On : 11 Jul 2023 11:58
 Operator : YP/AJ
 Sample : BSF0711S1
 Misc : 5.00G/5.00 ML DI WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF0711S1

Integration File: Calibration.e
 Quant Time: Jul 12 03:10:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB062623.M
 Quant Title :
 QLast Update : Mon Jun 26 14:31:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.790	279022	21.232 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.705	527244	28.046 ng/ml
2) t 2,2,4-Trimethylpentane	7.417	565967	26.915 ng/ml
3) t n-Heptane	7.748	201681	9.884 ng/ml
4) t Benzene	7.884	240864	9.829 ng/ml
6) t Toluene	10.618	691389	29.138 ng/ml
7) t Ethylbenzene	13.054	215790	9.589 ng/ml
8) t m-Xylene	13.186	428487	18.842 ng/ml
9) t O-Xylene	13.913	414680	18.291 ng/ml
10) t 1,2,4-Trimethylbenzene	16.188	406048	17.754 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB071123\
Data File : FB030152.D
Signal(s) : FID2B.CH
Acq On : 11 Jul 2023 11:58
Operator : YP/AJ
Sample : BSF0711S1
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF0711S1

Integration File: Calibration.e
Quant Time: Jul 12 03:10:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB062623.M
Quant Title :
QLast Update : Mon Jun 26 14:31:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

