

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB071318\
 Data File : FB016718.D
 Signal(s) : FID2B.CH
 Acq On : 13 Jul 2018 8:21
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jul 13 15:04:50 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070918.M
 Quant Title :
 QLast Update : Mon Jul 09 13:06:15 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.110	153787	17.639 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.961	374221	25.759 ng/ml
2) t	2,2,4-Trimethylpentane	7.725	384931	27.675 ng/ml
3) t	n-Heptane	8.064	128645	9.494 ng/ml
4) t	Benzene	8.203	167077	9.004 ng/ml
6) t	Toluene	10.958	399488	27.356 ng/ml
7) t	Ethylbenzene	13.407	133596	9.096 ng/ml
8) t	m-Xylene	13.540	282408	18.130 ng/ml
9) t	O-Xylene	14.272	302680	18.286 ng/ml
10) t	1,2,4-Trimethylbenzene	16.554	358917	18.352 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB071318\
Data File : FB016718.D
Signal(s) : FID2B.CH
Acq On : 13 Jul 2018 8:21
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jul 13 15:04:50 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070918.M
Quant Title :
QLast Update : Mon Jul 09 13:06:15 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.961 min
Delta R.T.: -0.005 min
Response: 374221
Conc: 25.76 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.725 min
Delta R.T.: -0.004 min
Response: 384931
Conc: 27.68 ng/ml




#3 n-Heptane

R.T.: 8.064 min
Delta R.T.: -0.006 min
Response: 128645
Conc: 9.49 ng/ml




#4 Benzene

R.T.: 8.203 min
Delta R.T.: -0.002 min
Response: 167077
Conc: 9.00 ng/ml




#5 AAA-TFT

R.T.: 9.110 min
Delta R.T.: 0.000 min
Response: 153787
Conc: 17.64 ng/ml




#6 Toluene

R.T.: 10.958 min
Delta R.T.: -0.001 min
Response: 399488
Conc: 27.36 ng/ml




#7 Ethylbenzene

R.T.: 13.407 min
Delta R.T.: -0.002 min
Response: 133596
Conc: 9.10 ng/ml




#8 m-Xylene

R.T.: 13.540 min
Delta R.T.: -0.002 min
Response: 282408
Conc: 18.13 ng/ml



#9 O-Xylene

R.T.: 14.272 min
Delta R.T.: -0.002 min
Response: 302680
Conc: 18.29 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.554 min
Delta R.T.: -0.002 min
Response: 358917
Conc: 18.35 ng/ml