

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB072419\
 Data File : FB021802.D
 Signal(s) : FID2B.CH
 Acq On : 24 Jul 2019 10:26
 Operator : DD/AJ
 Sample : K3618-07
 Misc : 5.06/5.00ml DI WATER
 ALS Vial : 4 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jul 25 01:05:08 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB071519.M
 Quant Title :
 QLast Update : Mon Jul 15 13:30:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.988	330301	24.204 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.852	789	0.029 ng/ml
2) t 2,2,4-Trimethylpentane	7.605	4377	0.166 ng/ml
4) t Benzene	8.091	3511	0.120 ng/ml
6) t Toluene	10.836	8626	0.419 ng/ml
8) t m-Xylene	13.404	7097	0.319 ng/ml
9) t O-Xylene	14.127	8580	0.373 ng/ml
10) t 1,2,4-Trimethylbenzene	16.419	9054	0.384 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB072419\
Data File : FB021802.D
Signal(s) : FID2B.CH
Acq On : 24 Jul 2019 10:26
Operator : DD/AJ
Sample : K3618-07
Misc : 5.06/5.00ml DI WATER
ALS Vial : 4 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jul 25 01:05:08 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB071519.M
Quant Title :
QLast Update : Mon Jul 15 13:30:07 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.852 min
Delta R.T.: -0.011 min
Response: 789
Conc: 0.03 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.605 min
Delta R.T.: -0.007 min
Response: 4377
Conc: 0.17 ng/ml




#4 Benzene

R.T.: 8.091 min
Delta R.T.: 0.005 min
Response: 3511
Conc: 0.12 ng/ml




#5 AAA-TFT

R.T.: 8.988 min
Delta R.T.: -0.004 min
Response: 330301
Conc: 24.20 ng/ml




#6 Toluene

R.T.: 10.836 min
Delta R.T.: 0.003 min
Response: 8626
Conc: 0.42 ng/ml




#8 m-Xylene

R.T.: 13.404 min
Delta R.T.: -0.007 min
Response: 7097
Conc: 0.32 ng/ml



#9 o-Xylene

R.T.: 14.127 min
Delta R.T.: -0.013 min
Response: 8580
Conc: 0.37 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.419 min
Delta R.T.: 0.000 min
Response: 9054
Conc: 0.38 ng/ml