

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB072518\
 Data File : FB016918.D
 Signal(s) : FID2B.CH
 Acq On : 25 Jul 2018 15:48
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jul 26 08:48:17 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070918.M
 Quant Title :
 QLast Update : Mon Jul 09 13:06:15 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.116	160361	18.393 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.965	428407	29.489 ng/ml
2) t 2,2,4-Trimethylpentane	7.733	424373	30.511 ng/ml
3) t n-Heptane	8.072	129225	9.537 ng/ml
4) t Benzene	8.212	172173	9.278 ng/ml
6) t Toluene	10.966	423047	28.970 ng/ml
7) t Ethylbenzene	13.416	139467	9.495 ng/ml
8) t m-Xylene	13.549	288430	18.517 ng/ml
9) t O-Xylene	14.281	304884	18.419 ng/ml
10) t 1,2,4-Trimethylbenzene	16.563	375391	19.195 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB072518\
Data File : FB016918.D
Signal(s) : FID2B.CH
Acq On : 25 Jul 2018 15:48
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jul 26 08:48:17 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070918.M
Quant Title :
QLast Update : Mon Jul 09 13:06:15 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.965 min
Delta R.T.: 0.000 min
Response: 428407
Conc: 29.49 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.733 min
Delta R.T.: 0.004 min
Response: 424373
Conc: 30.51 ng/ml




#3 n-Heptane

R.T.: 8.072 min
Delta R.T.: 0.002 min
Response: 129225
Conc: 9.54 ng/ml




#4 Benzene

R.T.: 8.212 min
Delta R.T.: 0.007 min
Response: 172173
Conc: 9.28 ng/ml




#5 AAA-TFT

R.T.: 9.116 min
Delta R.T.: 0.006 min
Response: 160361
Conc: 18.39 ng/ml




#6 Toluene

R.T.: 10.966 min
Delta R.T.: 0.007 min
Response: 423047
Conc: 28.97 ng/ml




#7 Ethylbenzene

R.T.: 13.416 min
Delta R.T.: 0.007 min
Response: 139467
Conc: 9.50 ng/ml




#8 m-Xylene

R.T.: 13.549 min
Delta R.T.: 0.007 min
Response: 288430
Conc: 18.52 ng/ml



#9 O-Xylene

R.T.: 14.281 min
Delta R.T.: 0.007 min
Response: 304884
Conc: 18.42 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.563 min
Delta R.T.: 0.007 min
Response: 375391
Conc: 19.19 ng/ml