

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB072618\
Data File : FB016949.D
Signal(s) : FID2B.CH
Acq On : 26 Jul 2018 14:58
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File: SAMPLE.e
Quant Time: Jul 27 08:07:12 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070918.M
Quant Title :
QLast Update : Mon Jul 09 13:06:15 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 t	2-Methylpentane	14.528	12.787 E3	12.0	93	0.00
2 t	2,2,4-Trimethylpentane	13.909	13.681 E3	1.6	104	0.00
3 t	n-Heptane	13.550	12.339 E3	8.9	95	0.00
4 t	Benzene	18.556	16.882 E3	9.0	96	0.00
5 s	AAA-TFT	8.719	8.755 E3	-0.4	107	0.00
6 t	Toluene	14.603	13.816 E3	5.4	99	0.00
7 t	Ethylbenzene	14.688	14.175 E3	3.5	98	0.00
8 t	m-Xylene	15.576	14.921 E3	4.2	96	0.00
9 t	O-Xylene	16.552	15.947 E3	3.7	97	0.00
10 t	1,2,4-Trimethylbenzene	19.557	18.878 E3	3.5	100	0.00

Evaluate Continuing Calibration Report - Not Founds

(#) = Out of Range

SPCC's out = 0 CCC's out = 0