

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB072618\
 Data File : FB016958.D
 Signal(s) : FID2B.CH
 Acq On : 26 Jul 2018 20:16
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Integration File: SAMPLE.e
 Quant Time: Jul 27 08:08:15 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070918.M
 Quant Title :
 QLast Update : Mon Jul 09 13:06:15 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 t	2-Methylpentane	14.528	13.624 E3	6.2	100	0.00
2 t	2,2,4-Trimethylpentane	13.909	14.257 E3	-2.5	108	0.00
3 t	n-Heptane	13.550	12.491 E3	7.8	97	0.00
4 t	Benzene	18.556	17.201 E3	7.3	98	0.00
5 s	AAA-TFT	8.719	9.362 E3	-7.4	114	0.00
6 t	Toluene	14.603	14.103 E3	3.4	101	0.00
7 t	Ethylbenzene	14.688	13.083 E3	10.9	91	0.00
8 t	m-Xylene	15.576	13.437 E3	13.7	87	0.00
9 t	O-Xylene	16.552	14.141 E3	14.6	86	0.00
10 t	1,2,4-Trimethylbenzene	19.557	17.532 E3	10.4	93	0.00

Evaluate Continuing Calibration Report - Not Founds

(#) = Out of Range

SPCC's out = 0 CCC's out = 0