

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB072618\
Data File : FB016965.D
Signal(s) : FID2B.CH
Acq On : 27 Jul 2018 00:19
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 26 Sample Multiplier: 1

Integration File: SAMPLE.e
Quant Time: Jul 27 08:09:04 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070918.M
Quant Title :
QLast Update : Mon Jul 09 13:06:15 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 t	2-Methylpentane	14.528	11.610 E3	20.1#	85	0.00
2 t	2,2,4-Trimethylpentane	13.909	12.367 E3	11.1	94	0.00
3 t	n-Heptane	13.550	10.800 E3	20.3#	84	0.00
4 t	Benzene	18.556	15.223 E3	18.0	87	0.00
5 s	AAA-TFT	8.719	8.705 E3	0.2	106	0.00
6 t	Toluene	14.603	12.541 E3	14.1	90	0.00
7 t	Ethylbenzene	14.688	13.003 E3	11.5	90	0.00
8 t	m-Xylene	15.576	13.508 E3	13.3	87	0.00
9 t	O-Xylene	16.552	14.259 E3	13.9	86	0.00
10 t	1,2,4-Trimethylbenzene	19.557	17.215 E3	12.0	91	0.00

Evaluate Continuing Calibration Report - Not Founds

(#) = Out of Range

SPCC's out = 0 CCC's out = 0