

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB072618\
 Data File : FB016949.D
 Signal(s) : FID2B.CH
 Acq On : 26 Jul 2018 14:58
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jul 27 08:07:12 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070918.M
 Quant Title :
 QLast Update : Mon Jul 09 13:06:15 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.114	175097	20.083 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.966	383624	26.407 ng/ml
2) t	2,2,4-Trimethylpentane	7.731	410428	29.508 ng/ml
3) t	n-Heptane	8.069	123386	9.106 ng/ml
4) t	Benzene	8.209	168818	9.098 ng/ml
6) t	Toluene	10.964	414474	28.383 ng/ml
7) t	Ethylbenzene	13.414	141751	9.651 ng/ml
8) t	m-Xylene	13.546	298414	19.158 ng/ml
9) t	O-Xylene	14.278	318946	19.269 ng/ml
10) t	1,2,4-Trimethylbenzene	16.560	377559	19.306 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB072618\
Data File : FB016949.D
Signal(s) : FID2B.CH
Acq On : 26 Jul 2018 14:58
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jul 27 08:07:12 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070918.M
Quant Title :
QLast Update : Mon Jul 09 13:06:15 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.966 min
Delta R.T.: 0.000 min
Response: 383624
Conc: 26.41 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.731 min
Delta R.T.: 0.002 min
Response: 410428
Conc: 29.51 ng/ml




#3 n-Heptane

R.T.: 8.069 min
Delta R.T.: 0.000 min
Response: 123386
Conc: 9.11 ng/ml




#4 Benzene

R.T.: 8.209 min
Delta R.T.: 0.004 min
Response: 168818
Conc: 9.10 ng/ml




#5 AAA-TFT

R.T.: 9.114 min
Delta R.T.: 0.004 min
Response: 175097
Conc: 20.08 ng/ml




#6 Toluene

R.T.: 10.964 min
Delta R.T.: 0.004 min
Response: 414474
Conc: 28.38 ng/ml




#7 Ethylbenzene

R.T.: 13.414 min
Delta R.T.: 0.005 min
Response: 141751
Conc: 9.65 ng/ml




#8 m-Xylene

R.T.: 13.546 min
Delta R.T.: 0.004 min
Response: 298414
Conc: 19.16 ng/ml



#9 O-Xylene

R.T.: 14.278 min
Delta R.T.: 0.004 min
Response: 318946
Conc: 19.27 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.560 min
Delta R.T.: 0.004 min
Response: 377559
Conc: 19.31 ng/ml