

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB073018\
 Data File : FB017013.D
 Signal(s) : FID2B.CH
 Acq On : 31 Jul 2018 5:57
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Jul 31 09:07:39 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB072718.M
 Quant Title :
 QLast Update : Fri Jul 27 13:23:50 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.112	170868	21.477 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.963	426305	29.967 ng/ml
2) t 2,2,4-Trimethylpentane	7.729	411560	30.003 ng/ml
3) t n-Heptane	8.067	129807	9.295 ng/ml
4) t Benzene	8.207	166739	9.519 ng/ml
6) t Toluene	10.961	424866	31.319 ng/ml
7) t Ethylbenzene	13.410	294230	21.440 ng/ml
8) t m-Xylene	13.543	462827	31.904 ng/ml
9) t O-Xylene	14.274	320141	20.808 ng/ml
10) t 1,2,4-Trimethylbenzene	16.557	358190	19.204 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB073018\
Data File : FB017013.D
Signal(s) : FID2B.CH
Acq On : 31 Jul 2018 5:57
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Jul 31 09:07:39 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB072718.M
Quant Title :
QLast Update : Fri Jul 27 13:23:50 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.963 min
Delta R.T.: -0.001 min
Response: 426305
Conc: 29.97 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.729 min
Delta R.T.: -0.002 min
Response: 411560
Conc: 30.00 ng/ml




#3 n-Heptane

R.T.: 8.067 min
Delta R.T.: -0.002 min
Response: 129807
Conc: 9.30 ng/ml



#4 Benzene


R.T.: 8.207 min
Delta R.T.: -0.002 min
Response: 166739
Conc: 9.52 ng/ml



#5 AAA-TFT


R.T.: 9.112 min
Delta R.T.: -0.002 min
Response: 170868
Conc: 21.48 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.961 min
Delta R.T.: -0.002 min
Response: 424866
Conc: 31.32 ng/ml




#7 Ethylbenzene

R.T.: 13.410 min
Delta R.T.: -0.002 min
Response: 294230
Conc: 21.44 ng/ml



#8 m-Xylene


R.T.: 13.543 min
Delta R.T.: -0.002 min
Response: 462827
Conc: 31.90 ng/ml



#9 O-Xylene

R.T.: 14.274 min
Delta R.T.: -0.002 min
Response: 320141
Conc: 20.81 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.557 min
Delta R.T.: -0.001 min
Response: 358190
Conc: 19.20 ng/ml